

REF 512-1	6x3mL	EXP 2023-10-13	LOT 30D6EEA	更新日期： 2021-02 1 / 2
------------------	-------	-----------------------	--------------------	------------------------

批号：30D6EEA				
项目\仪器\试剂方法	单位	均值	+ / - 2 SD	
(AFP)甲胎蛋白				
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	8.21	6.45-9.97	
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	7.92	6.16-9.68	
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	8.23	6.47-9.99	
HYBIOME EVERESYS Series HYBIOME \ Chemiluminescence Immunoassay (CLIA)	ng/mL	7.82	6.06-9.58	
(CA 125)糖类抗原 125				
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	35.6	28-43.2	
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	29.5	23.3-35.7	
HYBIOME EVERESYS Series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	23.9	18.7-29.1	
(CA 15-3)糖类抗原 15-3				
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	31.2	24.6-37.8	
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	32.4	25.2-39.6	
(CA 19-9)糖类抗原 19-9				
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/mL	31.4	25-37.8	
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	32.8	26-39.6	
(CA 27-29)糖类抗原 27-29				
Tosoh AIA1800 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	30.8	23.8-37.8	
(CA 72-4)糖类抗原 72-4				
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	6.98	5.54-8.42	
HYBIOME EVERESYS Series HYBIOME \ Chemiluminescence Immunoassay (CLIA)	IU/mL	11.2	9-13.4	
(CA242)糖类抗原 242				
HYBIOME EVERESYS Series HYBIOME \ Chemiluminescence Immunoassay (CLIA)	U/mL	11.9	9.5-14.3	
(CEA)癌胚抗原				
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.75	2.91-4.59	
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.65	2.09-3.21	
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3.23	2.55-3.91	
HYBIOME EVERESYS Series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.04	2.4-3.68	
(Cyfra 21-1)细胞角蛋白19片段				
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	4.21	3.27-5.15	
HYBIOME EVERESYS Series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.37	1.85-2.89	
(Ferritin)铁蛋白				
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	µg/L	31.2	24.6-37.8	
Beckman Access Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	µg/L	18.4	14.6-22.2	
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	µg/L	26.5	20.7-32.3	
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	µg/L	22700000	17900000-27500000	
HYBIOME EVERESYS Series Dedicated \ Chemiluminescence Immunoassay (CLIA)	µg/L	21.4	17.2-25.6	

REF 512-1	6x3mL	EXP 2023-10-13	LOT 30D6EEA	更新日期： 2021-02 2 / 2
------------------	-------	-----------------------	--------------------	------------------------

批号：30D6EEA				
项目\仪器\试剂\方法	单位	均值	+ / - 2 SD	
(F-PSA)游离前列腺特异性抗原				
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.31	0.23-0.39	
Beckman Access Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.47	0.39-0.55	
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.36	0.28-0.44	
(hCG)人绒毛膜促性腺激素				
Beckman Access Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/L	5.74	4.52-6.96	
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	6.62	5.2-8.04	
HYBIOME EVERESYS Series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/L	12.9	10.3-15.5	
(NSE)神经元特异性烯醇化酶				
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.27	1.01-1.53	
HYBIOME EVERESYS Series HYBIOME \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.51	1.99-3.03	
(PSA)前列腺特异性抗原				
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.46	1.14-1.78	
Beckman Access Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.57	1.23-1.91	
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.48	1.16-1.8	
HYBIOME EVERESYS Series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.52	1.2-1.84	
(β2M)β-2-微球蛋白				
HYBIOME EVERESYS Series Dedicated \ Chemiluminescence Immunoassay (CLIA)	mg/L	4.49	3.55-5.43	

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。