

REF 521-2	6x3mL	EXP 2024-04-04	LOT 31B51F8	更新日期： 2021-07
				1 / 2

批号：31B51F8				
项目\仪器\试剂\方法	单位	均值	+ / - 2 SD	
(Alb)白蛋白				
Roche Hitachi/Hitachi Systems Roche \ Bromocresol purple	g/L	33.2	25.8-40.6	
(ASO)抗链球菌溶血素 O 抗体				
Beckman AU series Beckman Coulter \ Immunoturbidimetry (ITA)	IU/mL	154	120-188	
Beckman Immage 800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	138	114-162	
Roche Hitachi/Hitachi Systems Wako \ Immunoturbidimetry (ITA)	IU/mL	154	120-188	
(C3)补体 C3				
Beckman AU series Beckman Coulter \ Immunoturbidimetry (ITA)	mg/dL	137	111-163	
Beckman Immage 800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	mg/dL	123	99-147	
(C4)补体 C4				
Beckman AU series Beckman Coulter \ Immunonephelometry (INA)	mg/dL	42	36-48	
Beckman Immage 800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	mg/dL	32.5	25.7-39.3	
(Cer)铜蓝蛋白				
Siemens BN II Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	g/L	0.321	0.257-0.385	
(CRP)C反应蛋白				
Beckman AU series Beckman Coulter \ Immunoturbidimetry (ITA)	mg/dL	2.63	1.99-3.27	
Beckman AU series KHB \ Immunoturbidimetry (ITA)	mg/dL	2.08	1.68-2.48	
Beckman Immage 800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	mg/dL	2.09	1.65-2.53	
(Hap (HPT))触珠蛋白 (结合珠蛋白)				
Siemens BN II Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	g/L	1.37	1.1-1.64	
(IgA)免疫球蛋白 IgA				
Beckman AU series Beckman Coulter \ Immunoturbidimetry (ITA)	g/L	2.04	1.58-2.5	
Beckman Immage 800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	g/L	2.11	1.65-2.57	
Roche Hitachi/Hitachi Systems Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	g/L	2.06	1.6-2.52	
(IgE)免疫球蛋白 IgE				
Beckman AU series Beckman Coulter \ Immunoturbidimetry (ITA)	IU/mL	39.7	32.3-47.1	
Beckman Immage 800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	38.2	30.6-45.8	
(IgG)免疫球蛋白 IgG				
Beckman AU series Beckman Coulter \ Immunoturbidimetry (ITA)	g/L	11.6	9-14.2	
Beckman Immage 800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	g/L	12.5	10-15	
Roche Hitachi/Hitachi Systems Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	g/L	12.7	10.2-15.2	
(IgM)免疫球蛋白 IgM				
Beckman AU series Beckman Coulter \ Immunoturbidimetry (ITA)	g/L	1.38	0.94-1.82	
Beckman Immage 800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	g/L	1.28	0.96-1.6	
Roche Hitachi/Hitachi Systems Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	g/L	1.25	0.93-1.57	
(Kappa Light Chain)κ轻链抗体				
Beckman Immage 800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	g/L	9.34	7.42-11.3	
Siemens BN II Dedicated \ Chemiluminescence Immunoassay (CLIA)	g/L	9.34	7.38-11.3	
(Lambda Light Chain)λ 轻链抗体				
Beckman Immage 800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	g/L	4.78	3.76-5.8	

REF 521-2	6x3mL	EXP 2024-04-04	LOT 31B51F8	更新日期： 2021-07 2 / 2
------------------	-------	-----------------------	--------------------	------------------------

批号：31B51F8				
项目\仪器\试剂方法	单位	均值	+ / - 2 SD	
Siemens BN II Dedicated \ Chemiluminescence Immunoassay (CLIA)	g/L	4.32	3.36-5.28	
(PAB)前白蛋白				
Beckman AU series Beckman Coulter \ Immunoturbidimetry (ITA)	g/L	0.265	0.209-0.321	
Beckman Immage 800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	g/L	0.377	0.309-0.445	
(RBP)视黄醇结合蛋白				
Siemens BN II Dedicated \ Chemiluminescence Immunoassay (CLIA)	mg/dL	4.43	3.51-5.35	
(RF)类风湿因子				
Beckman AU series Beckman Coulter \ Immunoturbidimetry (ITA)	IU/mL	62.6	50-75.2	
Beckman Immage 800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	59.4	47.8-71	
Roche Hitachi/Hitachi Systems Wako \ Immunoturbidimetry (ITA)	IU/mL	58.3	46.5-70.1	
(TRF)转铁蛋白				
Beckman Immage 800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	mg/dL	235	189-281	
(α-1-AG)α-1-酸性糖蛋白				
Beckman AU series Beckman Coulter \ Immunonephelometry (INA)	g/L	0.865	0.693-1.04	
(α-1-AT)α-1-抗胰蛋白酶				
Beckman AU series Beckman Coulter \ Immunoturbidimetry (ITA)	mg/dL	123	95-151	
(β2M)β-2-微球蛋白				
Beckman Immage 800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	mg/L	4.3	3.1-5.5	
Siemens BN II Dedicated \ Chemiluminescence Immunoassay (CLIA)	mg/L	3.6	2.8-4.4	

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。