

<b>REF</b> 513-2	6x3mL	<b>EXP</b> 2026-01-08	<b>LOT</b> 30F1CF5	更新日期： 2021-09 1 / 2
------------------	-------	-----------------------	--------------------	------------------------

批号：30F1CF5				
项目\仪器\试剂方法	单位	均值	+ / - 2 SD	
<b>(AFP)甲胎蛋白</b>				
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA )	ng/mL	79	60-98	
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	80.5	61.2-99.8	
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	84.4	64.1-105	
Siemens ADVIA Centaur XP Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	65.9	50.1-81.7	
<b>(CA 125)糖类抗原 125</b>				
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA )	IU/mL	230	175-285	
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA )	IU/mL	238	181-295	
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	170	129-211	
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	159	121-197	
Siemens ADVIA Centaur XP Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	182	138-226	
<b>(CA 15-3)糖类抗原 15-3</b>				
Siemens ADVIA Centaur XP Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	67.7	51.5-83.9	
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	53.1	40.4-65.8	
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	60.9	46.3-75.5	
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA )	IU/mL	62	47.1-76.9	
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA )	IU/mL	63.3	48.1-78.5	
<b>(CA 19-9)糖类抗原 19-9</b>				
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA )	IU/mL	241	183-299	
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA )	IU/mL	608	462-754	
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	123	93.5-153	
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	122	92.7-151	
Siemens ADVIA Centaur XP Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	239	182-296	
<b>(CA 50)糖类抗原 50</b>				
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA )	IU/mL	100	75.9-124	
<b>(CA 72-4)糖类抗原 72-4</b>				
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	13	9.89-16.1	
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	14.5	11-18	
<b>(CA242)糖类抗原 242</b>				
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA )	U/mL	35.1	26.7-43.5	
<b>(CEA)癌胚抗原</b>				
Siemens ADVIA Centaur XP Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	24.4	18.5-30.3	
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	36.2	27.5-44.9	
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	38.3	29.1-47.5	
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA )	ng/mL	26.7	20.3-33.1	
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA )	ng/mL	44	33.4-54.6	

<b>REF</b> 513-2	6x3mL	<b>EXP</b> 2026-01-08	<b>LOT</b> 30F1CF5	更新日期： 2021-09 2 / 2
------------------	-------	-----------------------	--------------------	------------------------

批号：30F1CF5				
项目\仪器\试剂\方法	单位	均值	+ / - 2 SD	
<b>(Cyfra 21-1)细胞角蛋白19片段</b>				
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	18.6	14.1-23.1	
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	19.9	15.1-24.7	
<b>(Ferritin)铁蛋白</b>				
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	µg/L	313	238-388	
Siemens ADVIA Centaur XP Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	µg/L	260	198-322	
Roche cobas c 501/502/701/702 Roche \ Immunoturbidimetry (ITA)	µg/L	318	242-394	
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA )	µg/L	391	297-485	
Beckman Access Dedicated \ Chemiluminescence Immunoassay (CLIA )	µg/L	221	168-274	
<b>(F-PSA)游离前列腺特异性抗原</b>				
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	4.54	3.45-5.63	
Beckman Access Dedicated \ Chemiluminescence Immunoassay (CLIA )	ng/mL	5.36	4.07-6.65	
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA )	ng/mL	4.9	3.72-6.08	
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	4	3.04-4.96	
Siemens ADVIA Centaur XP Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	4.53	3.44-5.62	
<b>(hCG)人绒毛膜促性腺激素</b>				
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	148	112-184	
Siemens ADVIA Centaur XP Dedicated \ Enzyme linked immunosorbent assays (ELISA)	IU/L	87	66.1-108	
<b>(HE4)人附睾蛋白4</b>				
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	256	195-317	
<b>(PSA)前列腺特异性抗原</b>				
Siemens ADVIA Centaur XP Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	4.53	3.44-5.62	
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	4	3.04-4.96	
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	21.4	16.3-26.5	
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA )	ng/mL	16.3	12.4-20.2	
Beckman Access Dedicated \ Chemiluminescence Immunoassay (CLIA )	ng/mL	18.2	13.8-22.6	
<b>(SCC)鳞状细胞癌抗原</b>				
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	9.99	7.59-12.4	
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA )	ng/mL	14.4	10.9-17.9	
<b>(TPA)组织多肽抗原</b>				
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA )	U/L	3550	2690-4400	
<b>(TPS)组织多肽特异抗原</b>				
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence Immunoassay (CLIA )	U/L	121	92-150	
<b>(β-hCG)人绒毛膜促性腺激素β亚基</b>				
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	161	122-200	

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。