

REF 510-1	6x3mL	EXP 2026-07-22	LOT 30A620C	更新日期： 2021-09 1 / 2
------------------	-------	-----------------------	--------------------	------------------------

批号：30A620C				
项目\仪器\试剂方法	单位	均值	+ / - 2 SD	
(AFP)甲胎蛋白				
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.34	3.3-5.38	
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	4.2	3.19-5.21	
Siemens ADVIA Centaur XP Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	3.2	2.43-3.97	
(CA 125)糖类抗原 125				
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	35	26.6-43.4	
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	29	22-36	
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	24.5	18.6-30.4	
Siemens ADVIA Centaur XP Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	23.7	18-29.4	
(CA 15-3)糖类抗原 15-3				
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	18	13.7-22.3	
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	17.1	13-21.2	
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	16.2	12.3-20.1	
Siemens ADVIA Centaur XP Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	16.5	12.5-20.5	
(CA 19-9)糖类抗原 19-9				
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	37	28.1-45.9	
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	24.1	18.3-29.9	
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	15.6	11.9-19.3	
Siemens ADVIA Centaur XP Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	24.7	18.8-30.6	
(CA 72-4)糖类抗原 72-4				
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	3.3	2.51-4.09	
(CEA)癌胚抗原				
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.3	2.51-4.09	
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.22	1.69-2.75	
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.3	1.75-2.85	
Siemens ADVIA Centaur XP Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	1.7	1.29-2.11	
(Cyfra 21-1)细胞角蛋白19片段				
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2	1.52-2.48	
(Ferritin)铁蛋白				
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	µg/L	52	39.5-64.5	
Beckman Access Dedicated \ Chemiluminescence Immunoassay (CLIA)	µg/L	34.2	26-42.4	
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	µg/L	62.2	47.3-77.1	
Siemens ADVIA Centaur XP Dedicated \ Chemiluminescence Immunoassay (CLIA)	µg/L	39	29.6-48.4	
(F-PSA)游离前列腺特异性抗原				
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.1	0.836-1.36	
Beckman Access Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.5	1.14-1.86	

REF 510-1	6x3mL	EXP 2026-07-22	LOT 30A620C	更新日期： 2021-09 2 / 2
------------------	-------	-----------------------	--------------------	------------------------

批号：30A620C				
项目\仪器\试剂方法	单位	均值	+ / - 2 SD	
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.96	0.73-1.19	
Siemens ADVIA Centaur XP Siemens \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.08	0.821-1.34	
(PSA)前列腺特异性抗原				
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.98	2.26-3.7	
Beckman Access Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.01	2.29-3.73	
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.97	2.26-3.68	
Siemens ADVIA Centaur XP Siemens \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.42	1.84-3	
(SCC)鳞状细胞癌抗原				
Abbott AxSYM Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.5	1.14-1.86	
(TPA)组织多肽抗原				
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA)	U/L	452	344-560	
(TPS)组织多肽特异抗原				
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence Immunoassay (CLIA)	U/L	4	3.04-4.96	
(β-hCG)人绒毛膜促性腺激素β亚基				
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	13.2	10-16.4	
Siemens ADVIA Centaur XP Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/L	10.7	8.13-13.3	

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。