

<b>REF</b> 513-1	6x3mL	<b>EXP</b> 2027-04-22	<b>LOT</b> 30EF64B	更新日期： 2022-06
				1 / 2

批号：30EF64B				
项目\仪器\试剂\方法	单位	均值	+ / - 2 SD	
<b>(AFP)甲胎蛋白</b>				
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	4.9	3.72-6.08	
<b>(CA 125)糖类抗原 125</b>				
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	IU/mL	41	31.2-50.8	
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	27.6	21-34.2	
Siemens ADVIA Centaur XP Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	31.1	23.6-38.6	
<b>(CA 15-3)糖类抗原 15-3</b>				
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	13.8	10.5-17.1	
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	IU/mL	14	10.6-17.4	
Siemens ADVIA Centaur XP Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	15.4	11.7-19.1	
<b>(CA 19-9)糖类抗原 19-9</b>				
Siemens ADVIA Centaur XP Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	22.9	17.4-28.4	
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	IU/mL	43	32.7-53.3	
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	10.8	8.21-13.4	
<b>(CA 72-4)糖类抗原 72-4</b>				
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	3.66	2.78-4.54	
<b>(CA242)糖类抗原 242</b>				
Siemens Immulite1000/2000 Siemens \ Chemiluminescence Immunoassay (CLIA )	U/mL	10.9	8.28-13.5	
<b>(CEA)癌胚抗原</b>				
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.8	2.13-3.47	
Siemens ADVIA Centaur XP Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	2.6	1.98-3.22	
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	4.1	3.12-5.08	
<b>(Cyfra 21-1)细胞角蛋白19片段</b>				
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3.5	2.66-4.34	
<b>(Ferritin)铁蛋白</b>				
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	µg/L	95.4	72.5-118	
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	µg/L	91	69.2-113	
<b>(F-PSA)游离前列腺特异性抗原</b>				
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	0.7	0.532-0.868	
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.61	0.464-0.756	
Siemens ADVIA Centaur XP Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	0.7	0.532-0.868	
<b>(hCG)人绒毛膜促性腺激素</b>				
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	15.2	11.6-18.8	
<b>(HE4)人附睾蛋白4</b>				
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	47.9	36.4-59.4	
<b>(PSA)前列腺特异性抗原</b>				
Siemens ADVIA Centaur XP Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	2.53	1.92-3.14	
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.76	2.1-3.42	
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	2.73	2.07-3.39	

<b>REF</b> 513-1	6x3mL	<b>EXP</b> 2027-04-22	<b>LOT</b> 30EF64B	更新日期： 2022-06 2 / 2
------------------	-------	-----------------------	--------------------	------------------------

批号：30EF64B			
项目\仪器\试剂\方法	单位	均值	+ / - 2 SD
<b>(SCC)鳞状细胞癌抗原</b>			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	1.5	1.14-1.86
<b>(TPA)组织多肽抗原</b>			
DiaSorin Liaison /XL Diasorin \ Chemiluminescence Immunoassay (CLIA )	U/L	269	204-334
<b>(TPS)组织多肽特异抗原</b>			
Siemens Immulite1000/2000 Siemens \ Chemiluminescence Immunoassay (CLIA )	U/L	17	12.9-21.1

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。