

<b>REF</b> 500-1	6x3mL	<b>EXP</b> 2025-05-04	<b>LOT</b> 2FB202B	更新日期: 2024-06
				1 / 3

本报告含以下项目的示范实验室数据

分析项目(Analyte)	分析项目(Analyte)	分析项目(Analyte)
(25-OH-Vit D)25羟基维生素D	(Cortisol)皮质醇	(C-P)C-肽
(E2)雌二醇	(FA)叶酸	(FSH)促卵泡刺激素
(FT3)游离T3	(FT4 )游离T4	(INS)胰岛素
(LH)促黄体生成素	(PRL)催/泌乳素	(PROG)孕酮
(PSA)前列腺特异性抗原	(TES)睾酮	(TSH)促甲状腺素
(TT3)总三碘甲状原氨酸	(TT4)总T4	(VB12)维生素B12
(β-hCG)人绒毛膜促性腺激素β亚基		

注: 此报告所提供的项目与数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来。此报告不可代替产品说明书。此报告仅供学习、参考之用。因所用技术、仪器和试剂的不同, 或因制造商检测方法的改变, 均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求, 实验室须遵循相关技术规范确立自己的均值和可接受范围。

<b>REF</b> 500-1	6x3mL	<b>EXP</b> 2025-05-04	<b>LOT</b> 2FB202B	更新日期: 2024-06
				2 / 3

批号: 2FB202B				
项目\仪器\试剂方法	单位	均值	+ / - 2 SD	
<b>(25-OH-Vit D)25羟基维生素D</b>				
Abbott ARCHITECT i series				
Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	11.3	8.58-14	
Abbott ARCHITECT i series				
Abbott \ Chemiluminescence Immunoassay (CLIA )	nmol/L	28.4	21.6-35.2	
<b>(Cortisol)皮质醇</b>				
Abbott ARCHITECT i series				
Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	31.3	23.8-38.8	
<b>(C-P)C-肽</b>				
Abbott ARCHITECT i series				
Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	0.932	0.708-1.16	
<b>(E2)雌二醇</b>				
Abbott ARCHITECT i series				
Abbott \ Chemiluminescence Immunoassay (CLIA )	pmol/L	258	196-320	
Autolumo A2000/A2000 PLUS				
Autobio \ Chemiluminescence Immunoassay (CLIA )	pmol/L	411	312-510	
<b>(FA)叶酸</b>				
Abbott ARCHITECT i series				
Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	0.202	0.154-0.25	
<b>(FSH)促卵泡刺激素</b>				
Abbott ARCHITECT i series				
Abbott \ Chemiluminescence Immunoassay (CLIA )	IU/L	4.02	3.05-4.99	
Autolumo A2000/A2000 PLUS				
Autobio \ Chemiluminescence Immunoassay (CLIA )	IU/L	5.83	4.43-7.23	
<b>(FT3)游离T3</b>				
Abbott ARCHITECT i series				
Abbott \ Chemiluminescence Immunoassay (CLIA )	pmol/L	4.15	3.15-5.15	
Autolumo A2000/A2000 PLUS				
Autobio \ Chemiluminescence Immunoassay (CLIA )	pmol/L	3.86	2.93-4.79	
<b>(FT4 )游离T4</b>				
Abbott ARCHITECT i series				
Abbott \ Chemiluminescence Immunoassay (CLIA )	pmol/L	7.84	5.96-9.72	
Autolumo A2000/A2000 PLUS				
Autobio \ Chemiluminescence Immunoassay (CLIA )	pmol/L	9.83	7.47-12.2	
<b>(INS)胰岛素</b>				
Abbott ARCHITECT i series				
Abbott \ Chemiluminescence Immunoassay (CLIA )	mIU/L	15.2	11.6-18.8	
<b>(LH)促黄体生成素</b>				
Abbott ARCHITECT i series				
Abbott \ Chemiluminescence Immunoassay (CLIA )	IU/L	1.13	0.858-1.4	
Autolumo A2000/A2000 PLUS				
Autobio \ Chemiluminescence Immunoassay (CLIA )	IU/L	3.86	2.93-4.79	
<b>(PRL)催/泌乳素</b>				
Abbott ARCHITECT i series				
Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	2.8	2.13-3.47	
Autolumo A2000/A2000 PLUS				
Autobio \ Chemiluminescence Immunoassay (CLIA )	ng/mL	3.56	2.71-4.41	
<b>(PROG)孕酮</b>				
Abbott ARCHITECT i series				
Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	1.62	1.23-2.01	
Autolumo A2000/A2000 PLUS				
Autobio \ Chemiluminescence Immunoassay (CLIA )	ng/mL	1.22	0.926-1.51	
<b>(PSA)前列腺特异性抗原</b>				
maglumi system				
Snibe \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	2.04	1.64-2.44	
<b>(TES)睾酮</b>				
Abbott ARCHITECT i series				
Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	0.668	0.508-0.828	
Autolumo A2000/A2000 PLUS				
Autobio \ Chemiluminescence Immunoassay (CLIA )	ng/mL	0.781	0.594-0.968	
<b>(TSH)促甲状腺素</b>				
Abbott ARCHITECT i series				
Abbott \ Chemiluminescence Immunoassay (CLIA )	mIU/L	0.029	0.021-0.037	
Autolumo A2000/A2000 PLUS				
Autobio \ Chemiluminescence Immunoassay (CLIA )	mIU/L	0.021	0.015-0.027	

注: 此报告所提供的项目与数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来。此报告不可代替产品说明书。此报告仅供学习、参考之用。因所用技术、仪器和试剂的不同, 或因制造商检测方法的改变, 均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求, 实验室须遵循相关技术规范确立自己的均值和可接受范围。

<b>REF</b> 500-1	6x3mL	<b>EXP</b> 2025-05-04	<b>LOT</b> 2FB202B	更新日期： 2024-06
				3 / 3

(TT3)总三碘甲状腺原氨酸			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	0.626	0.476-0.776
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA )	ng/mL	0.492	0.374-0.61
(TT4)总T4			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	nmol/L	24.2	18.4-30
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA )	nmol/L	37.3	28.4-46.2
(VB12)维生素B12			
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	pg/mL	163	124-202
( $\beta$ -hCG)人绒毛膜促性腺激素 $\beta$ 亚基			
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	3.94	3.14-4.74

注：此报告所提供的项目与数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来。此报告不可代替产品说明书。此报告仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。