

REF 510-1	6x3mL	EXP 2027-02-10	LOT 30A626C	更新日期: 2024-06
				1 / 2

本报告含以下项目的示范实验室数据

分析项目(Analyte)	分析项目(Analyte)	分析项目(Analyte)
(AFP)甲胎蛋白	(CA 125)糖类抗原 125	(CA 15-3)糖类抗原 15-3
(CA 19-9)糖类抗原 19-9	(CA 50)糖类抗原 50	(CA 72-4)糖类抗原 72-4
(CA242)糖类抗原 242	(CEA)癌胚抗原	(Cyfra 21-1)细胞角蛋白19片段
(Ferritin)铁蛋白	(F-PSA)游离前列腺特异性抗原	(HE4)人附睾蛋白4
(proGRP)胃泌素释放肽前体	(PSA)前列腺特异性抗原	(SCC)鳞状细胞癌抗原
(β -hCG)人绒毛膜促性腺激素 β 亚基		

注: 此报告所提供的项目与数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来。此报告不可代替产品说明书。此报告仅供学习、参考之用。因所用技术、仪器和试剂的不同, 或因制造商检测方法的改变, 均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求, 实验室须遵循相关技术规范确立自己的均值和可接受范围。

REF 510-1	6x3mL	EXP 2027-02-10	LOT 30A626C	更新日期： 2024-06
				2 / 2

批号： 30A626C				
项目\仪器\试剂方法	单位	均值	+ / - 2 SD	
(AFP)甲胎蛋白				
Abbott ARCHITECT c series Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.2	1.76-2.64	
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.87	1.49-2.25	
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA)	ng/mL	5.15	4.12-6.18	
(CA 125)糖类抗原 125				
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA)	IU/mL	30.3	24.2-36.4	
(CA 15-3)糖类抗原 15-3				
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA)	IU/mL	14.9	11.9-17.9	
(CA 19-9)糖类抗原 19-9				
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA)	IU/mL	10.5	8.41-12.6	
(CA 50)糖类抗原 50				
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	IU/mL	8.78	7.02-10.5	
(CA 72-4)糖类抗原 72-4				
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA)	IU/mL	3.39	2.71-4.07	
(CA242)糖类抗原 242				
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	U/mL	6.09	4.87-7.31	
(CEA)癌胚抗原				
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.89	2.31-3.47	
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.77	1.41-2.13	
Abbott ARCHITECT c series Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.89	1.51-2.27	
(Cyfra 21-1)细胞角蛋白19片段				
Abbott ARCHITECT c series Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.76	0.604-0.916	
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.82	0.66-0.98	
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.98	2.38-3.58	
(Ferritin)铁蛋白				
Roche cobas 4000/6000/8000 Roche \ Immunonephelometry (INA)	µg/L	65.3	52.2-78.4	
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	µg/L	89.6	71.7-108	
(F-PSA)游离前列腺特异性抗原				
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.676	0.541-0.811	
(HE4)人附睾蛋白4				
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	pmol/L	30.6	24.5-36.7	
Abbott ARCHITECT c series Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	33.5	26.8-40.2	
(proGRP)胃泌素释放肽前体				
Abbott ARCHITECT c series Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/L	192000000	153000000-230000000	
Abbott ARCHITECT i series Abbott \ Electro chemi luminescence (ECL)	ng/L	166000000	132000000-199000000	
(PSA)前列腺特异性抗原				
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.84	2.27-3.41	
(SCC)鳞状细胞癌抗原				
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.23	0.984-1.48	
(β-hCG)人绒毛膜促性腺激素β亚基				
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA)	IU/L	16.8	13.4-20.2	

注：此报告所提供的项目与数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来。此报告不可代替产品说明书。此报告仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。