

REF	500-1	6x3mL	EXP	2025-05-04	LOT	2FB202B	更新日期:	2024-08
								1 / 2

		批号: 2FB202B					
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数			
(25-OH-Vit D)25羟基维生素D							
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	11.3	12%	<5	<100		
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	nmol/L	28.4	12%	<5	<100		
(Cortisol)皮质醇							
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	31.3	12%	<5	<100		
(C-P)C-肽							
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.932	12%	<5	<100		
(E2)雌二醇							
OTHER OTHER \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	492	19.2%	<5	<100		
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	pmol/L	411	12%	<5	<100		
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	pmol/L	258	12%	<5	<100		
(FA)叶酸							
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.202	12%	<5	<100		
(FSH)促卵泡刺激素							
OTHER OTHER \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	5.46	8.9%	<5	<100		
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	IU/L	5.83	12%	<5	<100		
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/L	4.02	12%	<5	<100		
(FT3)游离T3							
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	pmol/L	4.15	12%	<5	<100		
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	pmol/L	3.86	12%	<5	<100		
(FT4)游离T4							
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	pmol/L	9.83	12%	<5	<100		
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	pmol/L	7.84	12%	<5	<100		
(INS)胰岛素							
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	mIU/L	15.2	12%	<5	<100		
(LH)促黄体生成素							
OTHER OTHER \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	3.3	6.2%	<5	<100		
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	IU/L	3.87	12%	<5	<100		
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/L	1.13	12%	<5	<100		
(PRL)催/泌乳素							
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.56	12%	<5	<100		
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.8	12%	<5	<100		
(PROG)孕酮							
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.22	12%	<5	<100		
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.62	12%	<5	<100		
(PSA)前列腺特异性抗原							
maglumi system Snibe \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	2.04	8.2%	<5	<100		
(TES)睾酮							
OTHER OTHER \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.012	6.4%	<5	<100		
Mindray CL-8000i Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	8.83	0%	<5	<100		
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.668	12%	<5	<100		
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.781	12%	<5	<100		
(TSH)促甲状腺素							
OTHER OTHER \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	0.246	6.2%	<5	<100		
Mindray CL-8000i Mindray \ Chemiluminescence Immunoassay (CLIA)	mIU/L	2.24	0%	<5	<100		
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	mIU/L	0.021	12%	<5	<100		
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	mIU/L	0.029	12%	<5	<100		

REF	500-1	6x3mL	EXP	2025-05-04	LOT	2FB202B	更新日期:	2024-08
								2 / 2

批号: 2FB202B					
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
(TT3)总三碘甲状原氨酸					
OTHER OTHER \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.002	7.4%	<5	<100
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.492	12%	<5	<100
ORTHO VISION Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	10.1	20%	<5	<100
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.564	8.9%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.626	12%	<5	<100
(TT4)总T4					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	18.8	12%	<5	<100
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	ng/mL	29	12%	<5	<100
(VB12)维生素B12					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	pg/mL	163	12%	<5	<100
(β-hCG)人绒毛膜促性腺激素β亚基					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	0.004	10.2%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。