

<b>REF</b>	500-3	6x3mL	<b>EXP</b>	2025-05-01	<b>LOT</b>	2FB6E4C	更新日期:	2024-08
								1 / 3

		批号: 2FB6E4C					
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数			
<b>(25-OH-Vit D)25羟基维生素D</b>							
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	34.6	12%	<5	<100		
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	nmol/L	86.6	12%	<5	<100		
<b>(Cortisol)皮质醇</b>							
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	430	35.6%	<5	<100		
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	388	6.6%	13	293		
<b>(C-P)C-肽</b>							
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	11.2	6.8%	6	136		
Abbott ARCHITECT i series Abbott \ CHEMILUMINESCENCE Immunoassay (CLIA )	ng/mL	8.66	9.8%	<5	<100		
<b>(E2)雌二醇</b>							
Ortho Vitros 3600/5600 OTHER \ Chemiluminescence Immunoassay (CLIA )	pmol/L	2228	15.7%	6	<100		
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	1462	4.9%	6	137		
Beckman Access Beckman Coulter \ Chemiluminescence enzyme immunoassay (CLEIA)	pmol/L	229	6%	5	112		
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence enzyme immunoassay (CLEIA)	pmol/L	1760	4.5%	13	300		
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA )	pmol/L	3238	12%	<5	<100		
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	pmol/L	1286	12%	<5	<100		
<b>(FA)叶酸</b>							
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	2.97	12%	<5	<100		
<b>(FSH)促卵泡刺激素</b>							
Ortho Vitros 3600/5600 OTHER \ Chemiluminescence Immunoassay (CLIA )	IU/L	57.2	4.7%	6	<100		
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	62.5	5.9%	6	136		
Beckman Access Beckman Coulter \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/L	7.51	6.2%	5	105		
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/L	82.7	6.3%	13	316		
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA )	IU/L	42	12%	<5	<100		
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	IU/L	37.4	12%	<5	<100		
<b>(FT3)游离T3</b>							
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	14.6	4.3%	6	136		
OTHER TELLGEN \ Chemiluminescence Immunoassay (CLIA )	pmol/L	21.9	7.9%	5	139		
Beckman Access Beckman Coulter \ Chemiluminescence enzyme immunoassay (CLEIA)	pmol/L	2.39	8.3%	5	112		
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	pmol/L	16.5	18.8%	<5	<100		
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence enzyme immunoassay (CLEIA)	pmol/L	9.94	5.8%	13	300		
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA )	pmol/L	26.1	12%	<5	<100		
GeteinBiotech MAGICL Series Getein Biotech \ Chemiluminescence Immunoassay (CLIA )	pmol/L	9.46	1.1%	<5	<100		
<b>(FT4 )游离T4</b>							
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	41.9	4.5%	6	136		
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	pmol/L	33.3	11.6%	<5	<100		
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence enzyme immunoassay (CLEIA)	pmol/L	36.5	3.7%	13	307		
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA )	pmol/L	40.5	12%	<5	<100		
GeteinBiotech MAGICL Series Getein Biotech \ Chemiluminescence Immunoassay (CLIA )	pmol/L	28.8	2.1%	<5	<100		
<b>(hCG)人绒毛膜促性腺激素</b>							
Ortho Vitros 3600/5600 OTHER \ Chemiluminescence Immunoassay (CLIA )	IU/L	690	4.5%	6	<100		
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	IU/L	508	12%	<5	<100		
Beckman Access Beckman Coulter \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/L	15.7	10%	<5	<100		

<b>REF</b>	500-3	6x3mL	<b>EXP</b>	2025-05-01	<b>LOT</b>	2FB6E4C	更新日期:	2024-08
								2 / 3

		批号: 2FB6E4C			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
<b>(INS)胰岛素</b>					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	mIU/L	89.2	16.3%	<5	<100
<b>(LH)促黄体生成素</b>					
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	56.1	5.2%	8	165
Ortho Vitros 3600/5600 OTHER \ Chemiluminescence Immunoassay (CLIA )	IU/L	63.8	6.6%	6	<100
Beckman Access Beckman Coulter \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/L	2.05	7.2%	5	123
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/L	48.8	3.1%	13	307
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA )	IU/L	48	12%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	IU/L	18.6	12%	<5	<100
<b>(PRL)催乳素</b>					
Ortho Vitros 3600/5600 OTHER \ Chemiluminescence Immunoassay (CLIA )	ng/mL	35.5	6.3%	6	<100
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	48.9	6.5%	6	136
Beckman Access Beckman Coulter \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	6.52	8.3%	5	104
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	35.8	3.9%	13	308
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA )	ng/mL	67.1	12%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	45.3	12%	<5	<100
<b>(PROG)孕酮</b>					
Ortho Vitros 3600/5600 OTHER \ Chemiluminescence Immunoassay (CLIA )	ng/mL	16.2	5.7%	6	<100
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	14.1	6%	6	137
Beckman Access Beckman Coulter \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	0.916	7.4%	5	117
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	14.6	8.5%	13	323
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA )	ng/mL	26.7	12%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	22.9	12%	<5	<100
<b>(TES)睾酮</b>					
Ortho Vitros 3600/5600 OTHER \ Chemiluminescence Immunoassay (CLIA )	ng/mL	13.5	6%	6	<100
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	8.51	7.4%	6	136
Beckman Access Beckman Coulter \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	0.507	8.7%	5	114
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	7.45	5.6%	13	302
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	7.79	12%	<5	<100
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA )	ng/mL	12	12%	<5	<100
<b>(TSH)促甲状腺素</b>					
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	mIU/L	26.1	5.4%	6	131
OTHER TELLGEN \ Chemiluminescence Immunoassay (CLIA )	mIU/L	34.6	3.7%	5	138
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	mIU/L	21.5	28.6%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence enzyme immunoassay (CLEIA)	mIU/L	27.8	6.3%	13	300
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA )	mIU/L	22.9	12%	<5	<100
<b>(TT3)总三碘甲状腺原氨酸</b>					
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA )	ng/mL	4.01	12%	<5	<100
GeteinBiotech MAGICL Series Getein Biotech \ Chemiluminescence Immunoassay (CLIA )	ng/mL	1.86	2.6%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	2.99	12%	<5	<100
<b>(TT4)总T4</b>					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	105	12%	<5	<100
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA )	ng/mL	126	12%	<5	<100

<b>REF</b>	500-3	6x3mL	<b>EXP</b>	2025-05-01	<b>LOT</b>	2FB6E4C	更新日期:	2024-08
								3 / 3

批号: 2FB6E4C					
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
GeteinBiotech MAGICL Series Getein Biotech \ Chemiluminescence Immunoassay (CLIA )	ng/mL	45.5	2.9%	<5	<100
<b>(VB12)维生素B12</b>					
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pg/mL	737	4.5%	7	137
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence enzyme immunoassay (CLEIA)	pg/mL	608	5.4%	13	315
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	pg/mL	729	12%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。