

REF	500-3	6x3mL	EXP	2025-10-20	LOT	2FB6E52	更新日期:	2024-08
								1 / 2

		批号: 2FB6E52					
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数			
(25-OH-Vit D)25羟基维生素D							
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	35	12%	<5	<100		
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	nmol/L	87.4	12%	<5	<100		
(Cortisol)皮质醇							
Lifotronic ecl Series Lifotronic \ Chemiluminescence Immunoassay (CLIA)	ng/mL	243	3.3%	<5	<100		
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	254	12%	<5	<100		
(C-P)C-肽							
Lifotronic ecl Series Lifotronic \ Chemiluminescence Immunoassay (CLIA)	ng/mL	8.37	3.4%	<5	<100		
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	8.11	12%	<5	<100		
(E2)雌二醇							
Ortho Vitros 3600/5600 OTHER \ Chemiluminescence Immunoassay (CLIA)	pmol/L	1948	4.2%	10	167		
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	pmol/L	1299	12%	<5	<100		
Lifotronic ecl Series Lifotronic \ Chemiluminescence Immunoassay (CLIA)	pmol/L	1347	3.4%	<5	<100		
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	pmol/L	3270	12%	<5	<100		
(FA)叶酸							
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3	12%	<5	<100		
Lifotronic ecl Series Lifotronic \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.52	9.2%	<5	<100		
(FSH)促卵泡刺激素							
Ortho Vitros 3600/5600 OTHER \ Chemiluminescence Immunoassay (CLIA)	IU/L	56.8	2.8%	10	179		
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/L	37.8	12%	<5	<100		
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	IU/L	42.4	12%	<5	<100		
Lifotronic ecl Series Lifotronic \ Chemiluminescence Immunoassay (CLIA)	IU/L	40.2	2.2%	<5	<100		
(FT3)游离T3							
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	pmol/L	26.4	12%	<5	<100		
Lifotronic ecl Series Lifotronic \ Chemiluminescence Immunoassay (CLIA)	pmol/L	23	4.4%	<5	<100		
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	pmol/L	20	12%	<5	<100		
(FT4)游离T4							
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	pmol/L	40.9	12%	<5	<100		
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	pmol/L	32.8	12%	<5	<100		
Lifotronic ecl Series Lifotronic \ Chemiluminescence Immunoassay (CLIA)	pmol/L	37.4	3%	<5	<100		
(hCG)人绒毛膜促性腺激素							
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	IU/L	321	12.7%	6	104		
Ortho Vitros 3600/5600 OTHER \ Chemiluminescence Immunoassay (CLIA)	IU/L	647	4.8%	10	174		
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/L	514	12%	<5	<100		
Lifotronic ecl Series Lifotronic \ Chemiluminescence Immunoassay (CLIA)	IU/L	522	1.6%	<5	<100		
(INS)胰岛素							
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	mIU/L	100	12%	<5	<100		
Lifotronic ecl Series Lifotronic \ Chemiluminescence Immunoassay (CLIA)	mIU/L	97.9	4.4%	<5	<100		
(LH)促黄体生成素							
Ortho Vitros 3600/5600 OTHER \ Chemiluminescence Immunoassay (CLIA)	IU/L	63.5	4.9%	10	171		
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/L	18.8	12%	<5	<100		
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	IU/L	48.5	12%	<5	<100		
Lifotronic ecl Series Lifotronic \ Chemiluminescence Immunoassay (CLIA)	IU/L	44.7	3.1%	<5	<100		
(PRL)催/泌乳素							
Ortho Vitros 3600/5600 OTHER \ Chemiluminescence Immunoassay (CLIA)	ng/mL	35.1	3.7%	10	137		
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	ng/mL	67.7	12%	<5	<100		
Lifotronic ecl Series Lifotronic \ Chemiluminescence Immunoassay (CLIA)	ng/mL	62	3.4%	<5	<100		

REF	500-3	6x3mL	EXP	2025-10-20	LOT	2FB6E52	更新日期:	2024-08
								2 / 2

		批号: 2FB6E52			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	45.8	12%	<5	<100
(PROG)孕酮					
Ortho Vitros 3600/5600 OTHER \ Chemiluminescence Immunoassay (CLIA)	ng/mL	14.8	5.4%	10	166
Lifotronic ecl Series Lifotronic \ Chemiluminescence Immunoassay (CLIA)	ng/mL	24.2	4%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	23.1	12%	<5	<100
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	ng/mL	26.9	12%	<5	<100
Beckman AU series Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	22.4	5.3%	<5	<100
(TES)睾酮					
Ortho Vitros 3600/5600 OTHER \ Chemiluminescence Immunoassay (CLIA)	ng/mL	12.4	4.1%	10	131
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	ng/mL	12.1	12%	<5	<100
Lifotronic ecl Series Lifotronic \ Chemiluminescence Immunoassay (CLIA)	ng/mL	10.5	4.5%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	7.86	12%	<5	<100
(TSH)促甲状腺素					
Lifotronic ecl Series Lifotronic \ Chemiluminescence Immunoassay (CLIA)	mIU/L	22.1	4.1%	<5	<100
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	mIU/L	23.1	12%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	mIU/L	16.1	12%	<5	<100
(TT3)总三碘甲状腺原氨酸					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.02	12%	<5	<100
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.06	12%	<5	<100
Lifotronic ecl Series Lifotronic \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.33	12.8%	<5	<100
(TT4)总T4					
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	ng/mL	127	12%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	106	12%	<5	<100
Lifotronic ecl Series Lifotronic \ Chemiluminescence Immunoassay (CLIA)	ng/mL	128	1.7%	<5	<100
(VB12)维生素B12					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	pg/mL	736	12%	<5	<100
Lifotronic ecl Series Lifotronic \ Chemiluminescence Immunoassay (CLIA)	pg/mL	735	0.9%	<5	<100

注: 此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来, 仅供学习、参考之用。因所用技术、仪器和试剂的不同, 或因制造商检测方法的改变, 均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求, 实验室须遵循相关技术规范确立自己的均值和可接受范围。