

REF	500-3	6x3mL	EXP	2026-09-04	LOT	2FB6EB5	更新日期:	2024-08
								1 / 2

		批号: 2FB6EB5			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
(25-OH-Vit D)25羟基维生素D					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	38.5	10.9%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	nmol/L	96.2	10.9%	<5	<100
(Cortisol)皮质醇					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	279	10.9%	<5	<100
(C-P)C-肽					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	8.92	10.9%	<5	<100
(E2)雌二醇					
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	pmol/L	3597	10.9%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	pmol/L	1429	10.9%	<5	<100
(FA)叶酸					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.3	10.9%	<5	<100
(FSH)促卵泡刺激素					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/L	41.5	10.9%	<5	<100
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	IU/L	46.6	10.9%	<5	<100
(FT3)游离T3					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	pmol/L	22	10.9%	<5	<100
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	pmol/L	29	10.9%	<5	<100
(FT4)游离T4					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	pmol/L	36	10.9%	<5	<100
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	pmol/L	45	10.9%	<5	<100
(hCG)人绒毛膜促性腺激素					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/L	565	10.9%	<5	<100
(INS)胰岛素					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	mIU/L	110	10.9%	<5	<100
(LH)促黄体生成素					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/L	20.7	10.9%	<5	<100
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	IU/L	53.4	10.9%	<5	<100
(PRL)催乳素					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	50.3	10.9%	<5	<100
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	ng/mL	74.5	10.9%	<5	<100
(PROG)孕酮					
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	ng/mL	29.6	10.9%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	25.4	10.9%	<5	<100
(TES)睾酮					
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	ng/mL	13.3	10.9%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	8.65	10.9%	<5	<100
(TSH)促甲状腺素					
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	mIU/L	25.4	10.9%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	mIU/L	17.8	10.9%	<5	<100
(TT3)总三碘甲状腺原氨酸					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.32	10.9%	<5	<100
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.46	10.9%	<5	<100
(TT4)总T4					
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	ng/mL	140	10.9%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	116	10.9%	<5	<100

报告类别： 汇总实验室数据报告



质控品名： 内分泌学非定值质控品

REF	500-3	6x3mL	EXP	2026-09-04	LOT	2FB6EB5	更新日期:	2024-08
								2 / 2

		批号: 2FB6EB5			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(VB12)维生素B12					
Abbott ARCHITECT i series					
Abbott \ Chemiluminescence Immunoassay (CLIA)					
	pg/mL	810	10.9%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。