

REF	510-1	6x3mL	EXP	2024-06-18	LOT	30A6143	更新日期:	2024-08
								1 / 3

批号: 30A6143						
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数		
(ADA)腺苷脱氨酶						
AB Sciex API Series AllergScreen \ Peroxidase	U/L	2	0%	<5	<100	
(AFP)甲胎蛋白						
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	4.21	2.9%	<5	<100	
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.05	3%	<5	<100	
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.31	12%	<5	<100	
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.04	12%	<5	<100	
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.14	2.8%	<5	<100	
(Alb)白蛋白						
Beckman AU series Beckman Coulter \ Bromocresol green	g/L	1.05	6.7%	<5	<100	
(ALP)碱性磷酸酶						
AB Sciex API Series ABX \ dry chemistry	U/L	11.2	0%	<5	<100	
(Anti -HAV IgM)抗-甲肝病毒 IgM						
maglumi system Snibe \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	5	0%	<5	<100	
(AST/ SGOT)天门冬氨酸氨基转移酶						
AB Sciex API Series ABX \ dry chemistry	U/L	11.2	0%	<5	<100	
(CA 125)糖类抗原 125						
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	IU/mL	54.4	7.6%	6	<100	
Roche cobas E601/602 Roche \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	9.27	93.6%	40	883	
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	23	1.6%	<5	<100	
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	31.7	4.1%	<5	<100	
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	IU/mL	18.8	12%	<5	<100	
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA)	IU/mL	26.8	12%	<5	<100	
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	34.3	3.3%	<5	<100	
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	27.6	0.1%	<5	<100	
Abbott Alinity i Series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	70.9	0.2%	<5	<100	
(CA 15-3)糖类抗原 15-3						
Roche cobas E601/602 Roche \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	10.1	41.8%	41	922	
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	15.8	4.5%	<5	<100	
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	14.8	4.5%	<5	<100	
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA)	IU/mL	15.8	12%	<5	<100	
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	17.5	3.5%	<5	<100	
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	IU/mL	11.8	12%	<5	<100	
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	22.1	0.1%	<5	<100	
(CA 19-9)糖类抗原 19-9						
Roche cobas E601/602 Roche \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	9.27	41%	41	905	
Mindray CL-2000i/1000i/6000i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	31.2	14.7%	<5	<100	
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	12.3	4.1%	<5	<100	
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	28.4	5.6%	<5	<100	
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	29.4	0.2%	<5	<100	
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	26.9	6.6%	<5	<100	
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	IU/mL	11.6	12%	<5	<100	
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA)	IU/mL	14.1	12%	<5	<100	
(CA 50)糖类抗原 50						
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	IU/mL	11.7	6.4%	<5	<100	

REF	510-1	6x3mL	EXP	2024-06-18	LOT	30A6143	更新日期:	2024-08
								2 / 3

		批号: 30A6143			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
(CA 72-4)糖类抗原 72-4					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	1.87	4.9%	<5	<100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	2.03	4.2%	<5	<100
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA)	IU/mL	2.51	12%	<5	<100
(CEA)癌胚抗原					
Roche cobas E601/602 Roche \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	5.43	4.5%	8	112
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.41	5.6%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.76	4.4%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3.72	1%	<5	<100
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.18	12%	<5	<100
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.46	12%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.79	4.1%	<5	<100
(Cyfra 21-1)细胞角蛋白19片段					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.02	5.8%	<5	<100
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.38	12%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.29	5.7%	<5	<100
(DGXN)地高辛					
ABX Pentra Series ImmunoCAP \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	1.56	0%	<5	<100
(Ferritin)铁蛋白					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	62.8	3.6%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	58.5	0%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	66.6	2.6%	<5	<100
Roche cobas 4000/6000/8000 Roche \ Immunonephelometry (INA)	ng/mL	67.8	12%	<5	<100
(F-PSA)游离前列腺特异性抗原					
Roche cobas E601/602 Roche \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	0.994	6.7%	8	112
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.898	2.1%	<5	<100
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.96	12.1%	<5	<100
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.85	12%	<5	<100
(GLU)葡萄糖					
AB Sciex API Series ABX \ dry chemistry	mmol/L	62.2	0%	<5	<100
(PSA)前列腺特异性抗原					
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.86	2.2%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.75	3.5%	<5	<100
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.71	12%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.57	2.5%	<5	<100
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.23	12%	<5	<100
(SCC)鳞状细胞癌抗原					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.991	27.6%	9	167
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.39	5.5%	<5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.95	3.2%	<5	<100
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.12	12%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.04	7.6%	<5	<100
(β-hCG)人绒毛膜促性腺激素β亚基					
Roche cobas E601/602 Roche \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	0.01	6.6%	8	112

REF	510-1	6x3mL	EXP	2024-06-18	LOT	30A6143	更新日期:	2024-08
								3 / 3

		批号: 30A6143			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	0.013	3.8%	<5	<100
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA)	IU/mL	0.013	11.9%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。