

REF	510-1	6x3mL	EXP	2026-01-08	LOT	30A6205	更新日期:	2024-08
								1 / 2

		批号: 30A6205			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
(AFP)甲胎蛋白					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.89	3.8%	7	202
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	4.86	9.5%	<5	<100
Ortho Vitros 3600/5600 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.85	8.3%	<5	<100
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.99	12%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.81	10%	<5	<100
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.24	7.5%	<5	<100
AB Sciex API Series ABX \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	4.25	0%	<5	<100
(CA 125)糖类抗原 125					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	34.7	5%	7	202
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	26.6	4.9%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	23.7	8.9%	<5	<100
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA)	IU/mL	26.3	10%	<5	<100
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	24.8	9.7%	<5	<100
AB Sciex API Series ABX \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	30.6	0%	<5	<100
(CA 15-3)糖类抗原 15-3					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	19.7	6.2%	7	204
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	15.8	8.9%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	15.4	10.4%	<5	<100
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA)	IU/mL	16.9	10%	<5	<100
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	14.3	9.1%	<5	<100
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	13.1	9.2%	<5	<100
(CA 19-9)糖类抗原 19-9					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	37.2	6.1%	7	200
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	37.7	7.7%	<5	<100
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	19.6	9.7%	<5	<100
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA)	IU/mL	14.5	10%	<5	<100
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	23.9	10.9%	<5	<100
Siemens Immulite1000/2000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	16.1	8.7%	<5	<100
AB Sciex API Series ABX \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	39.3	0%	<5	<100
(CA 50)糖类抗原 50					
AutoLumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	IU/mL	10.9	10%	<5	<100
(CA 72-4)糖类抗原 72-4					
Mindray BC-6000 serices Mindray \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	1.43	7%	<5	106
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	5.43	11.6%	<5	<100
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA)	IU/mL	2.92	12%	<5	<100
(CA242)糖类抗原 242					
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	U/mL	7.28	12%	<5	<100
(CEA)癌胚抗原					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.98	4.4%	7	202
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.15	9.8%	<5	<100
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	2.25	11.6%	<5	<100
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.51	12%	<5	<100
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.67	8.2%	<5	<100

REF	510-1	6x3mL	EXP	2026-01-08	LOT	30A6205	更新日期:	2024-08
								2 / 2

		批号: 30A6205					
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数			
Siemens Immulite1000/2000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.87	11.1%	<5	<100		
AB Sciex API Series ABX \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.44	0%	<5	<100		
(Cyfra 21-1)细胞角蛋白19片段							
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.05	7.1%	7	202		
Mindray BC-6000 serices Mindray \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	3.49	3.4%	<5	107		
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.95	12%	<5	<100		
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.89	9%	<5	<100		
(Ferritin)铁蛋白							
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	53	3.3%	7	201		
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	75.6	8.2%	<5	<100		
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	57.8	9%	<5	<100		
Roche cobas 4000/6000/8000 Roche \ Immunonephelometry (INA)	ng/mL	62	10%	<5	<100		
(F-PSA)游离前列腺特异性抗原							
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.05	3.4%	7	202		
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.05	12%	<5	<100		
Beckman Access Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.47	8.2%	<5	<100		
AB Sciex API Series ABX \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.7	0%	<5	<100		
(PSA)前列腺特异性抗原							
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.31	4.1%	7	203		
Beckman Access Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.79	11.1%	<5	<100		
Siemens Immulite1000/2000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.47	8.5%	<5	<100		
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.83	7.8%	<5	<100		
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	2.04	8.3%	<5	<100		
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.17	12%	<5	<100		
(SCC)鳞状细胞癌抗原							
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.55	4.6%	7	201		
OTHER Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.54	38.4%	29	584		
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.25	12%	<5	<100		
Abbott AxSYM Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.36	8.8%	<5	<100		
(TPA)组织多肽抗原							
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA)	U/L	322	9%	<5	<100		
(TPS)组织多肽特异抗原							
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence Immunoassay (CLIA)	U/L	90.4	10%	<5	<100		
(β-hCG)人绒毛膜促性腺激素β亚基							
Beckman Access Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	0.012	1%	<5	<100		
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA)	IU/mL	0.014	10%	<5	<100		

注: 此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来, 仅供学习、参考之用。因所用技术、仪器和试剂的不同, 或因制造商检测方法的改变, 均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求, 实验室须遵循相关技术规范确立自己的均值和可接受范围。