

REF	510-2	6x3mL	EXP	2024-06-18	LOT	30A8853	更新日期:	2024-08
								1 / 2

		批号: 30A8853			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(AFP)甲胎蛋白					
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	83.7	1.5%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	78	0.3%	<5	<100
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA)	ng/mL	79.1	12%	<5	<100
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	69.1	12%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	90.6	1.7%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	88	2.4%	<5	<100
(CA 125)糖类抗原 125					
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	185	1.1%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	72.1	0.1%	<5	<100
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	176	0.6%	15	370
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	IU/mL	150	12%	<5	<100
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA)	IU/mL	217	12%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	258	2.5%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	247	1.5%	<5	<100
(CA 15-3)糖类抗原 15-3					
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	73.8	3.3%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	53.5	0.1%	<5	<100
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	34.7	5.5%	16	427
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA)	IU/mL	71.7	12%	<5	<100
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	IU/mL	37.6	12%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	75	3.2%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	77.5	2.7%	<5	<100
(CA 19-9)糖类抗原 19-9					
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	67.4	5.1%	<5	<100
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	65.9	2.5%	16	426
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA)	IU/mL	74.2	12%	<5	<100
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	IU/mL	69.1	12%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	80.2	0%	<5	<100
(CA 72-4)糖类抗原 72-4					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	7.42	1.3%	<5	<100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	5.19	10.2%	<5	<100
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA)	IU/mL	8.18	12%	<5	<100
(CEA)癌胚抗原					
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	25.6	1.3%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	20.1	1.6%	<5	<100
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA)	ng/mL	24.6	12%	<5	<100
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	18.5	12%	<5	<100
(Cyfra 21-1)细胞角蛋白19片段					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	15.3	1.6%	<5	<100
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA)	ng/mL	18.3	11.9%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	10.8	2.8%	<5	<100

REF	510-2	6x3mL	EXP	2024-06-18	LOT	30A8853	更新日期：	2024-08
								2 / 2

		批号： 30A8853			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
(Ferritin)铁蛋白					
Roche cobas E601/602	ng/mL	288	0.1%	<5	<100
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)					
Roche cobas 4000/6000/8000	ng/mL	366	12%	<5	<100
Roche \ Immunonephelometry (INA)					
Abbott ARCHITECT i series	ng/mL	364	2.4%	<5	<100
Dedicated \ Chemiluminescence Immunoassay (CLIA)					
Abbott ARCHITECT i series	ng/mL	342	3.3%	<5	<100
Abbott \ Chemiluminescence Immunoassay (CLIA)					
(F-PSA)游离前列腺特异性抗原					
Roche cobas 4000/6000/8000	ng/mL	4.39	1.7%	<5	<100
Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)					
Abbott ARCHITECT i series	ng/mL	4.08	1.9%	<5	<100
Dedicated \ Chemiluminescence Immunoassay (CLIA)					
Mindray CL-2000i/1000i/6000i series	ng/mL	4.41	12%	<5	<100
Mindray \ Chemiluminescence Immunoassay (CLIA)					
Roche cobas E601/602	ng/mL	4.61	12%	<5	<100
Roche \ Chemiluminescence Immunoassay (CLIA)					
Abbott ARCHITECT i series	ng/mL	4.04	1.6%	<5	<100
Abbott \ Chemiluminescence Immunoassay (CLIA)					
(hCG)人绒毛膜促性腺激素					
Abbott Alinity i Series	IU/L	191	2.2%	14	340
Abbott \ Chemiluminescence enzyme immunoassay (CLEIA)					
(PSA)前列腺特异性抗原					
Roche cobas 4000/6000/8000	ng/mL	14.9	2%	<5	<100
Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)					
Mindray CL-2000i/1000i/6000i series	ng/mL	26.9	11%	18	456
Mindray \ Chemiluminescence enzyme immunoassay (CLEIA)					
Roche cobas E601/602	ng/mL	18.1	11.9%	<5	<100
Roche \ Chemiluminescence Immunoassay (CLIA)					
Mindray CL-2000i/1000i/6000i series	ng/mL	13.9	12%	<5	<100
Mindray \ Chemiluminescence Immunoassay (CLIA)					
Abbott ARCHITECT i series	ng/mL	14.7	4.1%	<5	<100
Dedicated \ Chemiluminescence Immunoassay (CLIA)					
Abbott ARCHITECT i series	ng/mL	16.8	3.7%	<5	<100
Abbott \ Chemiluminescence Immunoassay (CLIA)					
(SCC)鳞状细胞癌抗原					
Roche cobas E601/602	ng/mL	11.4	11.1%	9	137
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)					
Roche cobas E601/602	ng/mL	12.8	2.5%	<5	<100
Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)					
Mindray CL-2000i/1000i/6000i series	ng/mL	8.37	12%	<5	<100
Mindray \ Chemiluminescence Immunoassay (CLIA)					
Abbott ARCHITECT i series	ng/mL	16.2	4.8%	<5	<100
Dedicated \ Chemiluminescence Immunoassay (CLIA)					
(β2M)β-2-微球蛋白					
Mindray CL-2000i/1000i/6000i series	ng/mL	4364	1%	10	259
Mindray \ OTHER					
(β-hCG)人绒毛膜促性腺激素β亚基					
Roche cobas 4000/6000/8000	IU/mL	0.139	2.8%	<5	<100
Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)					
Roche cobas E601/602	IU/mL	0.144	12%	<5	<100
Roche \ Chemiluminescence Immunoassay (CLIA)					

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。