

REF	510-2	6x3mL	EXP	2027-02-10	LOT	30A897C	更新日期:	2024-08
								1 / 2

		批号: 30A897C			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
(AFP)甲胎蛋白					
maglumi system					
Snibe \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	54.3	5.8%	7	103
Abbott ARCHITECT i series					
Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	52.4	4%	67	1528
DiaSorin Liaison /XL					
Diasorin \ Chemiluminescence Immunoassay (CLIA)	ng/mL	48.4	1%	<5	<100
Roche cobas e 801					
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	59.9	2.8%	12	280
Roche cobas E601/602					
Roche \ Chemiluminescence Immunoassay (CLIA)	ng/mL	50.4	10%	<5	<100
Roche cobas E601/602					
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	50.6	0.7%	<5	<100
(CA 125)糖类抗原 125					
Abbott ARCHITECT i series					
Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	273	23%	72	1627
maglumi system					
Snibe \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	157	4.9%	7	105
DiaSorin Liaison /XL					
Diasorin \ Chemiluminescence Immunoassay (CLIA)	IU/mL	310	0.2%	<5	<100
Roche cobas e 801					
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	201	2.5%	15	342
Abbott ARCHITECT c series					
Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	90.4	85%	11	266
Roche cobas E601/602					
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	221	0.8%	<5	<100
Roche cobas E601/602					
Roche \ Chemiluminescence Immunoassay (CLIA)	IU/mL	218	10%	<5	<100
(CA 15-3)糖类抗原 15-3					
Abbott ARCHITECT i series					
Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	113	14.8%	73	1663
maglumi system					
Snibe \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	167	9.9%	7	107
DiaSorin Liaison /XL					
Diasorin \ Chemiluminescence Immunoassay (CLIA)	IU/mL	114	0.5%	<5	<100
Roche cobas e 801					
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	100	4.9%	15	265
Abbott ARCHITECT c series					
Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	83.2	58.6%	11	245
Roche cobas E601/602					
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	112	1.9%	<5	<100
Roche cobas E601/602					
Roche \ Chemiluminescence Immunoassay (CLIA)	IU/mL	112	10%	<5	<100
(CA 19-9)糖类抗原 19-9					
maglumi system					
Snibe \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	59.8	4.3%	7	<100
Abbott ARCHITECT i series					
Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	34.7	8.3%	5	119
Roche cobas e 801					
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	50.7	2.7%	14	286
Abbott ARCHITECT c series					
Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	36.5	11.2%	10	201
Roche cobas E601/602					
Roche \ Chemiluminescence Immunoassay (CLIA)	IU/mL	46.3	10%	<5	<100
Roche cobas E601/602					
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	50.3	2.4%	<5	<100
(CA 50)糖类抗原 50					
Autolumo A2000/A2000 PLUS					
Autobio \ Chemiluminescence Immunoassay (CLIA)	IU/mL	39.5	10%	<5	<100
(CA 72-4)糖类抗原 72-4					
Roche cobas e 801					
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	12.7	4.4%	12	<100
Roche cobas E601/602					
Roche \ Chemiluminescence Immunoassay (CLIA)	IU/mL	14.4	10%	<5	<100
(CA242)糖类抗原 242					
maglumi system					
Snibe \ Chemiluminescence Immunoassay (CLIA)	U/mL	16.9	10%	<5	<100
(CEA)癌胚抗原					
Abbott ARCHITECT i series					
Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	33.9	22%	70	1648
maglumi system					
Snibe \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	21.7	4.3%	7	104
DiaSorin Liaison /XL					
Diasorin \ Chemiluminescence Immunoassay (CLIA)	ng/mL	23.1	31.7%	5	<100
Roche cobas e 801					
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	22.1	4.6%	18	381
Abbott ARCHITECT c series					
Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	23.6	50.5%	11	249
Roche cobas E601/602					
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	20.9	0.9%	<5	<100
Roche cobas E601/602					
Roche \ Chemiluminescence Immunoassay (CLIA)	ng/mL	20.7	10%	<5	<100
(Cyfra 21-1)细胞角蛋白19片段					
Abbott ARCHITECT i series					
Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	16.2	6%	6	<100

REF	510-2	6x3mL	EXP	2027-02-10	LOT	30A897C	更新日期:	2024-08
								2 / 2

		批号: 30A897C					
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数			
Abbott ARCHITECT i series Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	13.4	4.6%	<5	<100		
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	13.3	3.1%	12	<100		
Abbott ARCHITECT c series Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	16.1	6.6%	10	210		
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	13.9	2.4%	<5	<100		
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA)	ng/mL	13.3	10%	<5	<100		
(Ferritin)铁蛋白							
DiaSorin Liaison /XL Diasorin \ Chemiluminescence Immunoassay (CLIA)	ng/mL	342	0.4%	5	<100		
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	327	4.3%	18	423		
Roche cobas 4000/6000/8000 Roche \ Immunonephelometry (INA)	ng/mL	332	10%	<5	<100		
(F-PSA)游离前列腺特异性抗原							
maglumi system Snibe \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	5.72	6.7%	7	108		
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	8.79	15.3%	59	1398		
DiaSorin Liaison /XL Diasorin \ Chemiluminescence Immunoassay (CLIA)	ng/mL	5.01	0.2%	<5	<100		
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	5.88	4.2%	15	359		
Abbott ARCHITECT c series Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	11.9	28.7%	11	234		
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA)	ng/mL	7.32	10%	<5	<100		
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	7.42	0.7%	<5	<100		
(hCG)人绒毛膜促性腺激素							
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	138	2.9%	<5	<100		
(HE4)人附睾蛋白4							
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	15.6	1.9%	12	<100		
Abbott ARCHITECT c series Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	440	23.1%	11	229		
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	pmol/L	352	27%	10	194		
(proGRP)胃泌素释放肽前体							
Abbott ARCHITECT c series Abbott \ Electro chemi luminescence (ECL)	ng/L	376724853	10.4%	9	165		
Abbott ARCHITECT i series Abbott \ Electro chemi luminescence (ECL)	ng/L	343950782	14%	<5	<100		
Abbott ARCHITECT i series Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/L	347000000	10%	<5	<100		
Abbott ARCHITECT c series Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/L	393000000	10%	<5	<100		
(PSA)前列腺特异性抗原							
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	16.6	6.9%	67	1552		
DiaSorin Liaison /XL Diasorin \ Chemiluminescence Immunoassay (CLIA)	ng/mL	17.8	4.2%	5	<100		
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	16.5	2.7%	15	290		
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	19.9	1%	<5	<100		
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA)	ng/mL	19.4	10%	<5	<100		
(SCC)鳞状细胞癌抗原							
Roche cobas e 801 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	10.2	3.7%	9	<100		
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	15.7	17.1%	58	1304		
Abbott ARCHITECT c series Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	7.96	8.3%	11	208		
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	10.3	10%	<5	<100		
(β-hCG)人绒毛膜促性腺激素β亚基							
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA)	IU/mL	0.134	10%	<5	<100		

注: 此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来, 仅供学习、参考之用。因所用技术、仪器和试剂的不同, 或因制造商检测方法的改变, 均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求, 实验室须遵循相关技术规范确立自己的均值和可接受范围。