

REF	512-1	6x3mL	EXP	2023-10-13	LOT	30D6EEA	更新日期:	2024-08
								1 / 2

		批号: 30D6EEA					
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数			
(AFP)甲胎蛋白							
HYBIOME EVERESYS Series HYBIOME \ Chemiluminescence Immunoassay (CLIA)	ng/mL	7.82	11.3%	<5	<100		
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	7.92	11.1%	<5	<100		
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	8.21	10.7%	<5	<100		
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	8.23	10.7%	<5	<100		
(CA 125)糖类抗原 125							
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	35.6	10.7%	<5	<100		
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	29.5	10.5%	<5	<100		
HYBIOME EVERESYS Series HYBIOME \ Chemiluminescence Immunoassay (CLIA)	IU/mL	23.9	10.9%	<5	<100		
(CA 15-3)糖类抗原 15-3							
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	32.4	11.1%	<5	<100		
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	31.2	10.6%	<5	<100		
(CA 19-9)糖类抗原 19-9							
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	32.8	10.4%	<5	<100		
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/mL	31.4	10.2%	<5	<100		
(CA 27-29)糖类抗原 27-29							
Tosoh AIA 900/1800 TOSOH \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	30.8	11.4%	<5	<100		
(CA 72-4)糖类抗原 72-4							
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	6.98	10.3%	<5	<100		
HYBIOME EVERESYS Series HYBIOME \ Chemiluminescence Immunoassay (CLIA)	IU/mL	11.2	9.8%	<5	<100		
(CA242)糖类抗原 242							
HYBIOME EVERESYS Series HYBIOME \ Chemiluminescence Immunoassay (CLIA)	U/mL	11.9	10.1%	<5	<100		
(CEA)癌胚抗原							
HYBIOME EVERESYS Series HYBIOME \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.04	10.5%	<5	<100		
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.65	10.6%	<5	<100		
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.75	11.2%	<5	<100		
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3.23	10.5%	<5	<100		
(Cyfra 21-1)细胞角蛋白19片段							
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	4.21	11.2%	<5	<100		
HYBIOME EVERESYS Series HYBIOME \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.37	11%	<5	<100		
(Ferritin)铁蛋白							
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	26.5	10.9%	<5	<100		
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	31.2	10.6%	<5	<100		
Siemens Immulite1000/2000 Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	22700000	10.6%	<5	<100		
Beckman Access Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	18.4	10.3%	<5	<100		
HYBIOME EVERESYS Series HYBIOME \ Chemiluminescence Immunoassay (CLIA)	ng/mL	21.4	9.8%	<5	<100		
(F-PSA)游离前列腺特异性抗原							
Beckman Access Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.47	8.5%	<5	<100		
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.36	11.1%	<5	<100		
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.31	12.9%	<5	<100		
(hCG)人绒毛膜促性腺激素							
Beckman Access Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	IU/L	5.74	10.6%	<5	<100		
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	6.62	10.7%	<5	<100		
HYBIOME EVERESYS Series HYBIOME \ Chemiluminescence Immunoassay (CLIA)	IU/L	12.9	10.1%	<5	<100		

REF	512-1	6x3mL	EXP	2023-10-13	LOT	30D6EEA	更新日期:	2024-08
								2 / 2

批号: 30D6EEA					
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(NSE)神经元特异性烯醇化酶					
HYBIOME EVERESYS Series HYBIOME \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.51	10.4%	<5	<100
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.27	10.2%	<5	<100
(PSA)前列腺特异性抗原					
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.48	10.8%	<5	<100
Beckman Access Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.57	10.8%	<5	<100
HYBIOME EVERESYS Series HYBIOME \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.52	10.5%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.46	11%	<5	<100
(β2M)β-2-微球蛋白					
HYBIOME EVERESYS Series HYBIOME \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4485	10.4%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。