

<b>REF</b>	512-1	6x3mL	<b>EXP</b>	2024-04-03	<b>LOT</b>	30D6F48	更新日期:	2024-08
								1 / 2

		批号: 30D6F48			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
<b>(AFP)甲胎蛋白</b>					
OTHER					
Dedicated \ OTHER	ng/mL	15.9	6.5%	9	<100
Roche cobas 4000/6000/8000					
Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	9.07	5.4%	7	195
Abbott ARCHITECT i series					
Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	9.77	2.1%	<5	<100
Roche cobas E601/602					
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	12.2	5.8%	23	580
Roche cobas 4000/6000/8000					
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	12.4	5.3%	14	282
Beckman UniCel Dxl 600/800					
Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	12.5	10%	<5	<100
<b>(CA 125)糖类抗原 125</b>					
OTHER					
Dedicated \ OTHER	IU/mL	38.9	7.1%	8	<100
Roche cobas 4000/6000/8000					
Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	22.4	4.1%	7	197
Roche cobas E601/602					
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	22.5	6.2%	22	489
Abbott ARCHITECT i series					
Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	39.3	3.5%	<5	<100
Roche cobas 4000/6000/8000					
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	24.3	8.7%	12	265
<b>(CA 15-3)糖类抗原 15-3</b>					
Roche cobas 4000/6000/8000					
Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	30.6	7.2%	7	201
Roche cobas E601/602					
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	28.4	8.6%	23	574
Roche cobas 4000/6000/8000					
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	30.2	4.1%	12	263
Abbott ARCHITECT i series					
Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	19.9	2.3%	<5	<100
<b>(CA 19-9)糖类抗原 19-9</b>					
OTHER					
Dedicated \ OTHER	IU/mL	13.7	8.2%	8	<100
Roche cobas 4000/6000/8000					
Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	19.4	6.5%	8	213
Roche cobas E601/602					
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	24.4	7.4%	7	139
Roche cobas 4000/6000/8000					
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	22.5	5.8%	<5	<100
Abbott ARCHITECT i series					
Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	33.1	2.7%	<5	<100
Roche cobas 4000/6000/8000					
Snibe \ Chemiluminescence Immunoassay (CLIA)	IU/mL	22.5	4.2%	10	256
maglumi system					
Snibe \ Chemiluminescence Immunoassay (CLIA)	IU/mL	22.9	0%	<5	<100
<b>(CA 72-4)糖类抗原 72-4</b>					
Roche cobas 4000/6000/8000					
Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	5.4	3.8%	7	198
Roche cobas 4000/6000/8000					
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	5.51	8.3%	14	303
Roche cobas E601/602					
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	6.54	10%	<5	<100
<b>(CA242)糖类抗原 242</b>					
OTHER					
Dedicated \ OTHER	U/mL	12.2	7.5%	8	<100
<b>(CEA)癌胚抗原</b>					
OTHER					
Dedicated \ OTHER	ng/mL	4.89	7.1%	9	<100
Roche cobas 4000/6000/8000					
Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	4.27	4.2%	7	197
Abbott ARCHITECT i series					
Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3.59	2.7%	<5	<100
Roche cobas E601/602					
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	4.15	5.3%	23	535
Roche cobas 4000/6000/8000					
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	4.16	3.6%	12	268
Beckman UniCel Dxl 600/800					
Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.01	10%	<5	<100
<b>(Cyfra 21-1)细胞角蛋白19片段</b>					
OTHER					
Dedicated \ OTHER	ng/mL	4.53	6.7%	9	<100
Roche cobas 4000/6000/8000					
Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	4.45	4.5%	7	195
maglumi system					
Snibe \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3.21	24%	<5	<100

<b>REF</b>	512-1	6x3mL	<b>EXP</b>	2024-04-03	<b>LOT</b>	30D6F48	更新日期：	2024-08
								2 / 2

		批号：30D6F48			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Abbott ARCHITECT i series Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3.31	7.1%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	4.36	4.4%	23	577
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	4.72	3.1%	10	258
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA )	ng/mL	3.88	21.8%	<5	<100
<b>(Ferritin)铁蛋白</b>					
Roche cobas c 501/502/701/702 Roche \ Immunoturbidimetry (ITA)	ng/mL	39	10%	<5	<100
<b>(F-PSA)游离前列腺特异性抗原</b>					
OTHER Dedicated \ OTHER	ng/mL	0.327	8.9%	8	<100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.33	8.2%	7	196
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.277	2.3%	10	255
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.276	10%	<5	<100
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )	ng/mL	0.37	10%	<5	<100
<b>(Fβ-hCG)游离β-绒毛膜促性腺激素</b>					
OTHER Dedicated \ OTHER	ng/mL	0.501	2.4%	9	<100
<b>(NSE)神经元特异性烯醇化酶</b>					
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.931	12.2%	8	216
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA )	ng/mL	0.504	3.4%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.12	10%	<5	<100
<b>(PSA)前列腺特异性抗原</b>					
OTHER Dedicated \ OTHER	ng/mL	1.85	7.7%	8	<100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.33	3.7%	7	195
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.5	9.4%	23	562
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.57	2.7%	10	256
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )	ng/mL	1.89	10%	<5	<100
<b>(SCC)鳞状细胞癌抗原</b>					
Abbott ARCHITECT i series Abbott \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	2.16	9.2%	<5	<100
<b>(Tg)甲状腺球蛋白</b>					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	6.39	10%	<5	<100
<b>(β2M)β-2-微球蛋白</b>					
Roche cobas c 501/502/701/702 Roche \ Immunoturbidimetry (ITA)	ng/mL	4420	10%	<5	<100
<b>(β-hCG)人绒毛膜促性腺激素β亚基</b>					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	0.004	10%	<5	<100
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )	IU/mL	0.006	10%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。