

<b>REF</b>	512-3	6x3mL	<b>EXP</b>	2024-04-04	<b>LOT</b>	30DBD67	更新日期:	2024-08
								1 / 2

		批号: 30DBD67					
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数			
<b>(AFP)甲胎蛋白</b>							
maglumi system	ng/mL	275	5.8%	9	161		
Snibe \ Electro-Chemiluminescence Immunoassay (ECLIA)							
Roche cobas 4000/6000/8000	ng/mL	187	3.3%	7	200		
Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)							
Abbott ARCHITECT i series	ng/mL	116	2.7%	<5	<100		
Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)							
OTHER	ng/mL	205	4.7%	19	158		
WanTai DRD \ Chemiluminescence Immunoassay (CLIA )							
Roche cobas E601/602	ng/mL	282	10%	<5	<100		
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)							
Beckman UniCel Dxl 600/800	ng/mL	272	10%	<5	<100		
Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )							
<b>(CA 125)糖类抗原 125</b>							
maglumi system	IU/mL	142	4%	9	158		
Snibe \ Electro-Chemiluminescence Immunoassay (ECLIA)							
Roche cobas 4000/6000/8000	IU/mL	142	3.1%	7	202		
Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)							
Abbott ARCHITECT i series	IU/mL	131	1.9%	<5	<100		
Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)							
OTHER	IU/mL	229	4.6%	19	143		
WanTai DRD \ Chemiluminescence Immunoassay (CLIA )							
Roche cobas E601/602	IU/mL	158	10%	<5	<100		
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)							
<b>(CA 15-3)糖类抗原 15-3</b>							
Roche cobas 4000/6000/8000	IU/mL	137	6.8%	7	206		
Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)							
OTHER	IU/mL	150	17.5%	20	171		
WanTai DRD \ Chemiluminescence Immunoassay (CLIA )							
Abbott ARCHITECT i series	IU/mL	59.8	3%	<5	<100		
Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)							
Roche cobas E601/602	IU/mL	134	10%	<5	<100		
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)							
<b>(CA 19-9)糖类抗原 19-9</b>							
maglumi system	IU/mL	156	4.7%	9	160		
Snibe \ Electro-Chemiluminescence Immunoassay (ECLIA)							
Roche cobas 4000/6000/8000	IU/mL	121	5.5%	8	217		
Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)							
Abbott ARCHITECT i series	IU/mL	222	6.7%	<5	<100		
Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)							
OTHER	IU/mL	194	6.3%	17	114		
WanTai DRD \ Chemiluminescence Immunoassay (CLIA )							
Roche cobas E601/602	IU/mL	155	10%	<5	<100		
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)							
<b>(CA 72-4)糖类抗原 72-4</b>							
Roche cobas 4000/6000/8000	IU/mL	52.9	4.6%	7	199		
Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)							
Roche cobas 4000/6000/8000	IU/mL	55.5	9.2%	6	110		
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)							
OTHER	IU/mL	65.5	7.6%	19	131		
WanTai DRD \ Chemiluminescence Immunoassay (CLIA )							
Roche cobas E601/602	IU/mL	63.1	10%	<5	<100		
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)							
<b>(CEA)癌胚抗原</b>							
maglumi system	ng/mL	33.6	4.4%	9	159		
Snibe \ Electro-Chemiluminescence Immunoassay (ECLIA)							
Roche cobas 4000/6000/8000	ng/mL	28.4	3.1%	7	202		
Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)							
Abbott ARCHITECT i series	ng/mL	34.2	3.6%	<5	<100		
Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)							
OTHER	ng/mL	36.3	9.8%	20	175		
WanTai DRD \ Chemiluminescence Immunoassay (CLIA )							
Roche cobas E601/602	ng/mL	32.6	10%	<5	<100		
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)							
Beckman UniCel Dxl 600/800	ng/mL	36.7	10%	<5	<100		
Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )							
<b>(Cyfra 21-1)细胞角蛋白19片段</b>							
maglumi system	ng/mL	80.7	5.4%	8	145		
Snibe \ Electro-Chemiluminescence Immunoassay (ECLIA)							
Roche cobas 4000/6000/8000	ng/mL	64.4	3.1%	7	199		
Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)							
Roche cobas 4000/6000/8000	ng/mL	65.9	2.7%	6	119		
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)							
OTHER	ng/mL	72.8	6.2%	20	166		
WanTai DRD \ Chemiluminescence Immunoassay (CLIA )							
Roche cobas E601/602	ng/mL	64.7	10%	<5	<100		
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)							

<b>REF</b>	512-3	6x3mL	<b>EXP</b>	2024-04-04	<b>LOT</b>	30DBD67	更新日期:	2024-08
								2 / 2

		批号: 30DBD67			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
<b>(Ferritin)铁蛋白</b>					
OTHER WanTai DRD \ Chemiluminescence Immunoassay (CLIA )	ng/mL	519	13.9%	7	<100
Roche cobas c 501/502/701/702 Roche \ Immunoturbidimetry (ITA)	ng/mL	850	10%	<5	<100
<b>(F-PSA)游离前列腺特异性抗原</b>					
maglumi system Snibe \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3.36	5%	9	161
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	4.26	4.2%	7	200
Abbott ARCHITECT i series Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	5.97	3%	<5	<100
OTHER WanTai DRD \ Chemiluminescence Immunoassay (CLIA )	ng/mL	3.47	6.6%	12	147
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )	ng/mL	5.21	10%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	4.09	10%	<5	<100
<b>(NSE)神经元特异性烯醇化酶</b>					
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	21.8	4.7%	8	220
Roche cobas 4000/6000/8000 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	23.7	4.8%	6	122
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	23.1	10%	<5	<100
<b>(PSA)前列腺特异性抗原</b>					
maglumi system Snibe \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.171	5.6%	9	163
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	17.9	3.1%	7	197
Abbott ARCHITECT i series Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	6.49	4.9%	<5	<100
OTHER WanTai DRD \ Chemiluminescence Immunoassay (CLIA )	ng/mL	0.245	11.5%	19	169
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )	ng/mL	27.7	10%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	24.8	10%	<5	<100
<b>(SCC)鳞状细胞癌抗原</b>					
Abbott ARCHITECT i series Abbott \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	8.21	5.4%	<5	<100
<b>(Tg)甲状腺球蛋白</b>					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	105	10%	<5	<100
<b>(β2M)β-2-微球蛋白</b>					
Roche cobas c 501/502/701/702 Roche \ Immunoturbidimetry (ITA)	ng/mL	28960	10%	<5	<100
<b>(β-hCG)人绒毛膜促性腺激素β亚基</b>					
OTHER WanTai DRD \ Chemiluminescence Immunoassay (CLIA )	IU/mL	0.201	7.3%	13	154
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	0.198	10%	<5	<100
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )	IU/mL	0.207	10%	<5	<100

注: 此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来, 仅供学习、参考之用。因所用技术、仪器和试剂的不同, 或因制造商检测方法的改变, 均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求, 实验室须遵循相关技术规范确立自己的均值和可接受范围。