

REF	513-1	6x3mL	EXP	2024-03-22	LOT	30EF520	更新日期:	2024-08
								1 / 2

		批号: 30EF520			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
(AFP)甲胎蛋白					
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.5	6.9%	9	213
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.58	5.9%	27	658
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3.59	5%	<5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.67	10.6%	<5	<100
(CA 125)糖类抗原 125					
Beckman AU series Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	24.3	2.3%	9	183
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	37.3	5.5%	27	544
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	28.8	15.1%	21	338
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	38.9	3.4%	<5	<100
DiaSorin Liaison /XL Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	32.1	4%	<5	<100
(CA 15-3)糖类抗原 15-3					
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	12	9.2%	9	193
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	15.1	13.6%	27	655
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	18.9	7.5%	20	333
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	20.3	4.5%	<5	<100
(CA 19-9)糖类抗原 19-9					
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	18.5	5.4%	9	216
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	13.1	10.3%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	43.7	9.6%	28	644
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	17	5.3%	<5	<100
Mindray CL-2000i/1000i/6000i series Mindray \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	47	14.1%	15	158
(CA 50)糖类抗原 50					
Mindray CL-2000i/1000i/6000i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	10.3	1.2%	<5	<100
(CA 72-4)糖类抗原 72-4					
Beckman AU series Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	2.17	9%	8	157
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	IU/mL	2.46	26.6%	6	136
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	1.8	11.4%	<5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	1.54	7.2%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	1.64	6.3%	10	220
(CA242)糖类抗原 242					
Mindray CL-2000i/1000i/6000i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	U/mL	9.99	1.1%	<5	<100
(CEA)癌胚抗原					
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.2	5.8%	9	202
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.59	11%	25	543
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.6	12.4%	20	332
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.53	2.8%	<5	<100
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.58	3.6%	<5	<100
(Cyfra 21-1)细胞角蛋白19片段					
Beckman AU series Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.38	10.5%	9	189
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.02	8.8%	6	136

REF	513-1	6x3mL	EXP	2024-03-22	LOT	30EF520	更新日期:	2024-08
								2 / 2

		批号: 30EF520			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.42	4.7%	5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.88	17.1%	21	327
Abbott ARCHITECT i series Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.47	3.8%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.08	9.1%	10	208
(Ferritin)铁蛋白					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	70	7.1%	8	144
(F-PSA)游离前列腺特异性抗原					
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.938	5.5%	9	216
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.1	10.1%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.693	5.3%	25	619
(HE4)人附睾蛋白4					
Beckman AU series Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	42.6	2%	9	191
Abbott ARCHITECT i series Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	28	16.7%	7	147
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	pmol/L	36.7	5.8%	6	120
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	34.7	4%	<5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	26.2	18.2%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	40.1	10.5%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	pmol/L	45.5	5.5%	11	199
(NSE)神经元特异性烯醇化酶					
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.932	3.2%	<5	<100
(PSA)前列腺特异性抗原					
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.27	11.8%	9	206
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.73	5%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.1	18.5%	28	663
(SCC)鳞状细胞癌抗原					
Beckman AU series Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.68	5.9%	9	176
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	ng/mL	0.782	11.9%	7	145
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.84	6.2%	<5	<100
Mindray CL-2000i/1000i/6000i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.54	3.4%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.35	8.2%	12	242
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.01	4.3%	<5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.79	1.7%	<5	<100
(β2M)β-2-微球蛋白					
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	790	4.4%	<5	<100

注: 此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来, 仅供学习、参考之用。因所用技术、仪器和试剂的不同, 或因制造商检测方法的改变, 均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求, 实验室须遵循相关技术规范确立自己的均值和可接受范围。