

REF	513-1	6x3mL	EXP	2026-10-07	LOT	30EF5EE	更新日期:	2024-08
								1 / 3

		批号: 30EF5EE					
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数			
(AFP)甲胎蛋白							
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	5.18	9.4%	29	514		
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.19	10%	<5	<100		
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	4.23	10.2%	<5	<100		
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	3.31	11.2%	<5	<100		
(CA 125)糖类抗原 125							
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	24.9	3%	<5	<100		
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	23.9	2.7%	<5	<100		
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	28.1	5.1%	29	496		
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	37.4	10.2%	<5	<100		
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	30.5	11.8%	<5	<100		
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	25.5	11%	<5	<100		
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	32.6	12%	<5	<100		
(CA 15-3)糖类抗原 15-3							
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	16.2	6.3%	<5	<100		
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	15.9	11.1%	31	517		
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	14	4%	<5	<100		
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	15.6	10.3%	<5	<100		
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	15.2	12%	<5	<100		
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	15.4	11.7%	<5	<100		
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	15.4	12%	<5	<100		
(CA 19-9)糖类抗原 19-9							
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	13.2	5.1%	<5	<100		
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	9.94	17.6%	30	497		
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	12.8	3.4%	<5	<100		
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	27.1	10.3%	<5	<100		
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	43.2	10%	<5	<100		
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	25.4	12%	<5	<100		
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	16.7	12%	<5	<100		
(CA 50)糖类抗原 50							
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	IU/mL	11.5	11.7%	<5	<100		
(CA 72-4)糖类抗原 72-4							
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	2.88	3.9%	<5	<100		
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	2.49	2.8%	<5	<100		
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	2.81	8.7%	29	513		
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	3.23	12%	<5	<100		
(CA242)糖类抗原 242							
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	U/mL	8.21	10%	<5	<100		
(CEA)癌胚抗原							
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.79	15.7%	29	524		

REF	513-1	6x3mL	EXP	2026-10-07	LOT	30EF5EE	更新日期:	2024-08
								2 / 3

		批号: 30EF5EE					
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数			
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	3.32	9.9%	<5	<100		
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.26	11.7%	<5	<100		
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	1.79	11.7%	<5	<100		
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.26	9.7%	<5	<100		
(Cyfra 21-1)细胞角蛋白19片段							
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.21	4.4%	<5	<100		
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.93	10.4%	<5	<100		
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.52	5.9%	29	514		
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.33	12.4%	<5	<100		
(Ferritin)铁蛋白							
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	38.2	10%	<5	<100		
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	62.8	10.2%	<5	<100		
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	51.6	12.1%	<5	<100		
Beckman Access Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	32.4	10.5%	<5	<100		
Roche cobas c 501/502/701/702 Roche \ Immunoturbidimetry (ITA)	ng/mL	61.5	10.1%	<5	<100		
(F-PSA)游离前列腺特异性抗原							
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.623	27.1%	31	540		
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	1.09	12%	<5	<100		
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.98	9.2%	<5	<100		
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.31	10.1%	<5	<100		
Beckman Access Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.56	12%	<5	<100		
(hCG)人绒毛膜促性腺激素							
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	14.8	9%	<5	<100		
Siemens ADVIA Centaur XP/XPT Dedicated \ Enzyme linked immunosorbent assays (ELISA)	IU/L	9.47	12.3%	<5	<100		
(HE4)人附睾蛋白4							
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	35.9	2.5%	<5	<100		
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	53.7	6%	28	485		
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	38.4	9.9%	<5	<100		
(PSA)前列腺特异性抗原							
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	2.44	13.6%	29	532		
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	1.09	12%	<5	<100		
Beckman Access Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3.62	12%	<5	<100		
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	2.73	12.3%	<5	<100		
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.98	9.2%	<5	<100		
(SCC)鳞状细胞癌抗原							
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	1.15	9.2%	29	500		
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	1.52	11.8%	<5	<100		
(TPA)组织多肽抗原							
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA)	U/L	311	12%	<5	<100		
(TPS)组织多肽特异抗原							
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence Immunoassay (CLIA)	U/L	4.78	12.6%	<5	<100		

REF	513-1	6x3mL	EXP	2026-10-07	LOT	30EF5EE	更新日期:	2024-08
								3 / 3

		批号: 30EF5EE			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(β-hCG)人绒毛膜促性腺激素β亚基					
Roche cobas E601/602					
Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	0.015	11.9%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。