

REF	513-2	6x3mL	EXP	2024-03-22	LOT	30F1C30	更新日期:	2024-08
								1 / 3

		批号: 30F1C30			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
(AFP)甲胎蛋白					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	62.9	29.1%	29	695
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	52.1	12.9%	<5	<100
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	107	7.5%	12	275
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	106	6.1%	<5	<100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	77.9	3.7%	<5	<100
(CA 125)糖类抗原 125					
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	169	13.3%	7	105
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	288	7.3%	32	602
Beckman AU series Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	202	4%	11	243
DiaSorin Liaison /XL Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	309	2.7%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	350	2.1%	<5	<100
(CA 15-3)糖类抗原 15-3					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	117	8.4%	30	703
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	81.3	10.7%	25	404
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	38.2	8.4%	11	257
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	116	5%	<5	<100
(CA 19-9)糖类抗原 19-9					
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	90.2	6.2%	7	103
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	443	12.4%	30	690
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	163	7.4%	12	280
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	107	4.1%	<5	<100
(CA 50)糖类抗原 50					
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	IU/mL	66.6	5.8%	9	175
Mindray CL-2000i/1000i/6000i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	70.4	1.1%	<5	<100
maglumi system Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	48.7	1.6%	<5	<100
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	IU/mL	68.1	12%	<5	<100
maglumi system Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	49.7	6.7%	<5	<100
(CA 72-4)糖类抗原 72-4					
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	IU/mL	11.9	6.5%	9	174
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	5.39	5.7%	7	101
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	5.15	13.6%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	4.99	5.3%	10	216
Beckman AU series Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	5.17	7.8%	10	213
(CA242)糖类抗原 242					
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	U/mL	35.3	4.8%	6	120
maglumi system Dedicated \ Chemiluminescence Immunoassay (CLIA)	U/mL	40.5	21.1%	5	114
Mindray CL-2000i/1000i/6000i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	U/mL	56.8	1.3%	<5	<100
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	U/mL	56.6	12%	<5	<100
maglumi system Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	U/mL	24	2.4%	<5	<100

REF	513-2	6x3mL	EXP	2024-03-22	LOT	30F1C30	更新日期:	2024-08
								2 / 3

		批号: 30F1C30			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(CEA)癌胚抗原					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	51.1	30.5%	28	581
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	23.5	15.4%	<5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	29.2	9.5%	<5	<100
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	48.8	6.5%	12	262
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	29.9	2.8%	<5	<100
(cPSA)复合前列腺特异性抗原					
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/L	8366	3.8%	<5	<100
(Cyfra 21-1)细胞角蛋白19片段					
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	ng/mL	17.7	5%	9	174
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	13.7	4.5%	5	<100
Abbott ARCHITECT i series Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	14.1	2.5%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	11.3	4.3%	<5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	11.5	11.6%	13	182
Beckman AU series Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	11.6	3.7%	12	237
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	10.5	4.6%	10	204
(Ferritin)铁蛋白					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	398	8.9%	9	185
(F-PSA)游离前列腺特异性抗原					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	6.85	27.3%	28	662
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	4.59	6.3%	21	338
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	5.5	6.8%	12	277
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.62	2.2%	<5	<100
(HE4)人附睾蛋白4					
Abbott ARCHITECT i series Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	274	5%	9	187
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	pmol/L	412	8.6%	6	120
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	pmol/L	362	3.5%	<5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	311	12%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	318	13.4%	<5	<100
Beckman AU series Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	320	2.3%	11	252
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	pmol/L	260	5.4%	11	200
(NSE)神经元特异性烯醇化酶					
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	0.845	1.6%	<5	<100
(PSA)前列腺特异性抗原					
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	15	10.1%	35	682
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	16.1	13.7%	21	337
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	14.5	5.9%	12	265
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	15.7	2.6%	<5	<100
(SCC)鳞状细胞癌抗原					
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	ng/mL	13.5	6.9%	9	174
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	ng/mL	17	5.8%	<5	<100
Mindray CL-2000i/1000i/6000i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	14.2	1.7%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	16.3	7.3%	11	242

REF	513-2	6x3mL	EXP	2024-03-22	LOT	30F1C30	更新日期：	2024-08 3 / 3
------------	-------	-------	------------	------------	------------	---------	-------	------------------

		批号：30F1C30			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Beckman AU series Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	11.9	8.4%	11	235
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	14.3	11.9%	<5	<100
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	14.2	1.2%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	17.3	4.3%	<5	<100
(β2M)β-2-微球蛋白					
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	941	5.3%	<5	<100
(β-hCG)人绒毛膜促性腺激素β亚基					
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	0.208	1.1%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。