

<b>REF</b>	513-2	6x3mL	<b>EXP</b>	2025-02-05	<b>LOT</b>	30F1C92	更新日期:	2024-08
								1 / 3

		批号: 30F1C92			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
<b>(AFP)甲胎蛋白</b>					
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA )	ng/mL	83.3	5.7%	9	206
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	67.9	3.3%	14	248
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	80.9	10%	<5	<100
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	60.9	10%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	68.6	11%	<5	<100
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )	ng/mL	80.4	10%	<5	<100
<b>(CA 125)糖类抗原 125</b>					
Beckman AU series Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	166	3.1%	8	183
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	IU/mL	273	5.4%	15	281
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	186	11%	<5	<100
DiaSorin Liaison /XL Diasorin \ Chemiluminescence Immunoassay (CLIA )	IU/mL	273	10%	<5	<100
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	227	10%	<5	<100
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	176	10%	<5	<100
<b>(CA 15-3)糖类抗原 15-3</b>					
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA )	IU/mL	30.6	5.6%	9	201
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	IU/mL	109	12.9%	15	293
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	69.2	10%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	67.3	11%	<5	<100
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	96.4	10%	<5	<100
DiaSorin Liaison /XL Diasorin \ Chemiluminescence Immunoassay (CLIA )	IU/mL	94.5	10%	<5	<100
<b>(CA 19-9)糖类抗原 19-9</b>					
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA )	IU/mL	117	6.1%	9	190
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	IU/mL	367	8%	15	273
DiaSorin Liaison /XL Diasorin \ Chemiluminescence Immunoassay (CLIA )	IU/mL	177	10%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	74.3	11%	<5	<100
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	88	10%	<5	<100
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	179	10%	<5	<100
<b>(CA 50)糖类抗原 50</b>					
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA )	IU/mL	50.2	6.2%	11	218
<b>(CA 72-4)糖类抗原 72-4</b>					
Beckman AU series Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	4.38	4.8%	6	124
Roche cobas E601/602 Roche \ Chemiluminescence Immunoassay (CLIA )	IU/mL	4.58	4.7%	5	<100
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA )	IU/mL	18.8	31.6%	10	191
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	13.7	10%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	9.19	11%	<5	<100
<b>(CA242)糖类抗原 242</b>					
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA )	U/mL	25.6	5.9%	11	218
Manual GLD \ Enzyme linked immunosorbent assays (ELISA)	U/mL	40.1	10%	<5	<100
<b>(CEA)癌胚抗原</b>					
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA )	ng/mL	28.6	5.2%	9	207
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	44.4	5.3%	15	280
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	24.2	10%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	37.2	11%	<5	<100

<b>REF</b>	513-2	6x3mL	<b>EXP</b>	2025-02-05	<b>LOT</b>	30F1C92	更新日期:	2024-08
								2 / 3

		批号: 30F1C92					
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数			
Beckman UniCel DxI 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )	ng/mL	29.8	10%	<5	<100		
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	42.4	10%	<5	<100		
<b>(Cyfra 21-1)细胞角蛋白19片段</b>							
Beckman AU series Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA) maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA )	ng/mL	10.2	2.3%	9	191		
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	9.63	15.8%	5	<100		
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	9.72	4.1%	<5	105		
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	12.2	10%	<5	<100		
<b>(Ferritin)铁蛋白</b>							
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	375	5.5%	15	270		
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	291	10%	<5	<100		
Beckman Access Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )	ng/mL	247	10%	<5	<100		
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	280	10%	<5	<100		
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	336	11%	<5	<100		
<b>(F-PSA)游离前列腺特异性抗原</b>							
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA )	ng/mL	5.35	4.7%	9	204		
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	3.94	7%	15	280		
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	4.72	10%	<5	<100		
Beckman Access Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )	ng/mL	6.85	10%	<5	<100		
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	4.43	11%	<5	<100		
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	4.78	10%	<5	<100		
<b>(hCG)人绒毛膜促性腺激素</b>							
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	147	10%	<5	<100		
Siemens ADVIA Centaur XP/XPT Siemens \ Enzyme linked immunosorbent assays (ELISA)	IU/L	99	10%	<5	<100		
<b>(HE4)人附睾蛋白4</b>							
Beckman AU series Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	211	2.7%	7	128		
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	pmol/L	177	6.1%	<5	<100		
Abbott ARCHITECT i series Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	178	7.2%	11	220		
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	272	10%	<5	<100		
<b>(PSA)前列腺特异性抗原</b>							
Beckman AU series Dedicated \ Chemiluminescence Immunoassay (CLIA )	ng/mL	12.9	6.6%	9	202		
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	12.1	15.7%	15	284		
Beckman Access Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )	ng/mL	19.5	10%	<5	<100		
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	16.5	11%	<5	<100		
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	18.3	10%	<5	<100		
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	15	10%	<5	<100		
<b>(SCC)鳞状细胞癌抗原</b>							
Beckman AU series Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA) maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA )	ng/mL	7.8	3.5%	9	193		
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	14.5	7.3%	7	115		
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	12.7	8%	5	125		
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	8.52	11%	<5	<100		
<b>(TPA)组织多肽抗原</b>							
DiaSorin Liaison /XL Diasorin \ Chemiluminescence Immunoassay (CLIA )	U/L	3894	10%	<5	<100		

<b>REF</b>	513-2	6x3mL	<b>EXP</b>	2025-02-05	<b>LOT</b>	30F1C92	更新日期：	2024-08
								3 / 3

		批号：30F1C92			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
<b>(TPS)组织多肽特异抗原</b>					
Manual OTHER \ Enzyme linked immunosorbent assays (ELISA)	U/L	103	10%	<5	<100
<b>(β-hCG)人绒毛膜促性腺激素β亚基</b>					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	0.141	9.9%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。