

<b>REF</b>	513-2	6x3mL	<b>EXP</b>	2025-10-02	<b>LOT</b>	30F1C9A	更新日期:	2024-08
								1 / 3

		批号: 30F1C9A				
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数		
<b>(AFP)甲胎蛋白</b>						
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA )	ng/mL	71.6	4.4%	19	293	
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	68.1	3.8%	15	349	
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	81.5	9.9%	<5	<100	
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	69.4	10.4%	<5	<100	
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )	ng/mL	82.6	9.7%	<5	<100	
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	59.6	10.2%	<5	<100	
<b>(CA 125)糖类抗原 125</b>						
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA )	IU/mL	178	4.1%	18	293	
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	IU/mL	278	4.5%	16	374	
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	218	10.4%	<5	<100	
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	181	9.7%	<5	<100	
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	189	10.8%	<5	<100	
DiaSorin Liaison /XL Diasorin \ Chemiluminescence Immunoassay (CLIA )	IU/mL	274	10%	<5	<100	
<b>(CA 15-3)糖类抗原 15-3</b>						
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA )	IU/mL	27.4	10.2%	19	294	
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	IU/mL	110	5.8%	16	387	
DiaSorin Liaison /XL Diasorin \ Chemiluminescence Immunoassay (CLIA )	IU/mL	96.5	9.8%	<5	<100	
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	68.4	10.8%	<5	<100	
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	70.6	9.8%	<5	<100	
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	98.3	9.8%	<5	<100	
<b>(CA 19-9)糖类抗原 19-9</b>						
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA )	IU/mL	127	6.4%	19	335	
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	IU/mL	380	8.1%	16	361	
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	180	9.9%	<5	<100	
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	75.9	10.8%	<5	<100	
DiaSorin Liaison /XL Diasorin \ Chemiluminescence Immunoassay (CLIA )	IU/mL	182	9.7%	<5	<100	
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	89.6	9.8%	<5	<100	
<b>(CA 50)糖类抗原 50</b>						
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA )	IU/mL	52.5	6%	15	343	
<b>(CA 72-4)糖类抗原 72-4</b>						
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	5.44	26.8%	21	195	
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA )	IU/mL	6.51	4.2%	19	208	
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA )	IU/mL	7.92	63.9%	15	336	
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	12.6	10.9%	<5	<100	
<b>(CA242)糖类抗原 242</b>						
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA )	U/mL	39.7	1.9%	11	150	
Manual GLD \ Enzyme linked immunosorbent assays (ELISA)	U/mL	39.8	10.1%	<5	<100	
<b>(CEA)癌胚抗原</b>						
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA )	ng/mL	11.5	2.9%	<5	<100	
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	45.3	5.4%	16	381	
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	43.5	9.7%	<5	<100	
Beckman UniCel Dxl 600/800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )	ng/mL	30.2	9.9%	<5	<100	
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	25.8	9.4%	<5	<100	

<b>REF</b>	513-2	6x3mL	<b>EXP</b>	2025-10-02	<b>LOT</b>	30F1C9A	更新日期:	2024-08
								2 / 3

		批号: 30F1C9A					
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数			
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	36.4	11.2%	<5	<100		
<b>(Cyfra 21-1)细胞角蛋白19片段</b>							
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	10.1	5.9%	21	216		
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	8.24	15.8%	16	337		
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	12.6	9.7%	<5	<100		
<b>(Ferritin)铁蛋白</b>							
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	305	7.2%	21	198		
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA )	ng/mL	298	5.3%	19	328		
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	380	5.7%	16	380		
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	296	9.5%	<5	<100		
Beckman Access Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )	ng/mL	258	9.6%	<5	<100		
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	309	9.4%	<5	<100		
<b>(F-PSA)游离前列腺特异性抗原</b>							
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA )	ng/mL	4.34	3.4%	19	339		
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	3.99	5.6%	16	377		
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	4.88	9.7%	<5	<100		
Beckman Access Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )	ng/mL	5.96	11.5%	<5	<100		
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	4.96	9.6%	<5	<100		
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	4.56	10.7%	<5	<100		
<b>(HbA1c (%))糖化血红蛋白A1c (%)</b>							
MEDCONN MQ-2000PT HuaChenBio \ High performance liquid chromatography (HPLC)	%	9.4	0%	<5	<100		
<b>(hCG)人绒毛膜促性腺激素</b>							
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	149	9.9%	<5	<100		
Siemens ADVIA Centaur XP/XPT Siemens \ Enzyme linked immunosorbent assays (ELISA)	IU/L	103	9.6%	<5	<100		
<b>(HE4)人附睾蛋白4</b>							
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	194	3.7%	20	173		
Abbott ARCHITECT i series Abbott \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	172	7%	15	342		
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA )	pmol/L	239	3.1%	15	217		
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	268	10.1%	<5	<100		
<b>(PSA)前列腺特异性抗原</b>							
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA )	ng/mL	14.5	2.7%	<5	<100		
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	10.9	21.8%	16	395		
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	15.9	9.4%	<5	<100		
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	18.2	10.1%	<5	<100		
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	17.2	10.6%	<5	<100		
Beckman Access Beckman Coulter \ Chemiluminescence Immunoassay (CLIA )	ng/mL	18.6	10.5%	<5	<100		
<b>(SCC)鳞状细胞癌抗原</b>							
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	7.56	8.8%	21	198		
Mindray CL-2000i/1000i/6000i series Mindray \ Chemiluminescence Immunoassay (CLIA )	ng/mL	9.46	3.2%	19	224		
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA )	ng/mL	12.6	7.1%	16	381		
<b>(TPA)组织多肽抗原</b>							
DiaSorin Liaison /XL Diasorin \ Chemiluminescence Immunoassay (CLIA )	U/L	3768	10.3%	<5	<100		
<b>(TPS)组织多肽特异抗原</b>							
Manual OTHER \ Enzyme linked immunosorbent assays (ELISA)	U/L	99.6	10.3%	<5	<100		

<b>REF</b>	513-2	6x3mL	<b>EXP</b>	2025-10-02	<b>LOT</b>	30F1C9A	更新日期:	2024-08
								3 / 3

		批号: 30F1C9A			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
<b>(Urea)尿素</b>					
Roche cobas c 501/502/701/702 Roche \ Urease colorimetry	mmol/L	12.3	0%	<5	<100
<b>(β-hCG)人绒毛膜促性腺激素β亚基</b>					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	0.153	9.2%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。