

REF	513-2	6x3mL	EXP	2026-10-07	LOT	30F1CFE	更新日期:	2024-08
								1 / 3

		批号: 30F1CFE			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
(AFP)甲胎蛋白					
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	72.7	6.7%	6	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	54.1	14.2%	29	516
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	65.9	12%	<5	<100
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	79	12%	<5	<100
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	84.4	12%	<5	<100
(CA 125)糖类抗原 125					
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	173	4.5%	6	<100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	161	2.5%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	232	10%	29	510
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	182	12%	<5	<100
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	230	12%	<5	<100
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	159	12%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	238	12%	<5	<100
(CA 15-3)糖类抗原 15-3					
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	58.7	5.4%	6	<100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	58.3	7.2%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	111	17.1%	31	513
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	63.3	12%	<5	<100
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	53.1	12%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	62	12%	<5	<100
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	67.7	12%	<5	<100
(CA 19-9)糖类抗原 19-9					
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	104	14.5%	6	<100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	114	3%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	65	29.4%	30	511
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	122	12%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	608	12%	<5	<100
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	239	12%	<5	<100
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	241	12%	<5	<100
(CA 50)糖类抗原 50					
Autolumo A2000/A2000 PLUS Autobio \ Chemiluminescence Immunoassay (CLIA)	IU/mL	100	12%	<5	<100
(CA 72-4)糖类抗原 72-4					
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	14.1	4.1%	6	<100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	11.8	3.5%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	11.4	9.3%	29	516
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	14.5	12%	<5	<100
(CA242)糖类抗原 242					
maglumi system Snibe \ Chemiluminescence Immunoassay (CLIA)	U/mL	35.1	12%	<5	<100

REF	513-2	6x3mL	EXP	2026-10-07	LOT	30F1CFE	更新日期:	2024-08
								2 / 3

		批号: 30F1CFE			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
(CEA)癌胚抗原					
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	39.5	3.5%	6	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	42	9%	29	507
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	44	12%	<5	<100
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	24.4	12%	<5	<100
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	36.2	12%	<5	<100
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	26.7	12%	<5	<100
(Cyfra 21-1)细胞角蛋白19片段					
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	16.4	3.8%	6	<100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	17.4	2.6%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	14.5	13.3%	29	534
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	18.6	12%	<5	<100
(Ferritin)铁蛋白					
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	313	12%	<5	<100
Roche cobas c 501/502/701/702 Roche \ Immunoturbidimetry (ITA)	ng/mL	318	12%	<5	<100
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	260	12%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	391	12%	<5	<100
Beckman Access Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	221	12%	<5	<100
(F-PSA)游离前列腺特异性抗原					
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	4.45	4.3%	6	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	5.73	8.7%	31	537
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	4	12%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.9	12%	<5	<100
Beckman Access Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	5.36	12%	<5	<100
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	4.53	12%	<5	<100
(hCG)人绒毛膜促性腺激素					
Siemens ADVIA Centaur XP/XPT Dedicated \ Enzyme linked immunosorbent assays (ELISA)	IU/L	87	12%	<5	<100
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	148	12%	<5	<100
(HE4)人附睾蛋白4					
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	233	4.8%	6	<100
Roche cobas 4000/6000/8000 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	232	2.5%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	289	4.9%	28	484
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	256	12%	<5	<100
(PSA)前列腺特异性抗原					
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	13.3	4.3%	6	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	16.9	10.5%	29	532
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	4	12%	<5	<100
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	4.53	12%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	16.3	12%	<5	<100

REF	513-2	6x3mL	EXP	2026-10-07	LOT	30F1CFE	更新日期:	2024-08
								3 / 3

批号: 30F1CFE					
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Beckman Access Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	18.2	12%	<5	<100
(SCC)鳞状细胞癌抗原					
Roche cobas E601/602 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	9.13	5.2%	<5	<100
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	10.4	6.4%	29	518
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	14.4	12%	<5	<100
(TPA)组织多肽抗原					
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA)	U/L	3556	12%	<5	<100
(TPS)组织多肽特异抗原					
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence Immunoassay (CLIA)	U/L	121	12%	<5	<100
(β-hCG)人绒毛膜促性腺激素β亚基					
Roche cobas E601/602 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	0.161	12%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。