

REF	513-2	6x3mL	EXP	2027-11-26	LOT	30F1D64	更新日期:	2024-08
								1 / 4

		批号: 30F1D64			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
(AFP)甲胎蛋白					
Abbott Alinity i Series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	51.9	2.8%	12	306
Beckman UniCel Dxl 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	99.2	9.7%	<5	<100
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	52.6	10.1%	<5	<100
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	103	10.7%	<5	<100
Mindray CL-8000i Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	71.4	10%	<5	<100
HYBIOME EVERESYS Series HYBIOME \ Chemiluminescence Immunoassay (CLIA)	ng/mL	108	10%	<5	<100
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	116	9.5%	<5	<100
(CA 125)糖类抗原 125					
Abbott Alinity i Series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	321	3.4%	11	257
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	328	9.8%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	354	11.8%	<5	<100
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	391	9.7%	<5	<100
HYBIOME EVERESYS Series HYBIOME \ Chemiluminescence Immunoassay (CLIA)	IU/mL	281	10%	<5	<100
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	246	11.3%	<5	<100
Mindray CL-8000i Mindray \ Chemiluminescence Immunoassay (CLIA)	IU/mL	196	10%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	429	9.8%	<5	<100
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	266	9.8%	<5	<100
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	254	11.8%	<5	<100
(CA 15-3)糖类抗原 15-3					
Abbott Alinity i Series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	118	5.3%	12	316
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	95.9	9.7%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	119	9.3%	<5	<100
Mindray CL-8000i Mindray \ Chemiluminescence Immunoassay (CLIA)	IU/mL	30.7	10%	<5	<100
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	107	10.3%	<5	<100
HYBIOME EVERESYS Series HYBIOME \ Chemiluminescence Immunoassay (CLIA)	IU/mL	114	10%	<5	<100
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	118	11.6%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	125	11.4%	<5	<100
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	110	10%	<5	<100
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	109	11.3%	<5	<100
(CA 19-9)糖类抗原 19-9					
Abbott Alinity i Series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	496	6.1%	12	286
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	65.8	12.2%	<5	<100
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	99.1	8.4%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	404	9.4%	<5	<100
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	182	11.3%	<5	<100
HYBIOME EVERESYS Series HYBIOME \ Chemiluminescence Immunoassay (CLIA)	IU/mL	124	10%	<5	<100
Mindray CL-8000i Mindray \ Chemiluminescence Immunoassay (CLIA)	IU/mL	82.6	10%	<5	<100
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	IU/mL	188	8.5%	<5	<100
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	199	8%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	IU/mL	439	10%	<5	<100

REF	513-2	6x3mL	EXP	2027-11-26	LOT	30F1D64	更新日期:	2024-08
								2 / 4

		批号: 30F1D64					
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数			
(CA 50)糖类抗原 50							
HYBIOME EVERESYS Series HYBIOME \ Chemiluminescence Immunoassay (CLIA)	IU/mL	156	10%	<5	<100		
Mindray CL-8000i Mindray \ Chemiluminescence Immunoassay (CLIA)	IU/mL	64.8	10%	<5	<100		
(CA 72-4)糖类抗原 72-4							
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	12.4	9.7%	<5	<100		
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	16.3	12.4%	<5	<100		
Maccura i3000 MAKER \ Chemiluminescence Immunoassay (CLIA)	IU/mL	16.2	3.3%	<5	<100		
HYBIOME EVERESYS Series HYBIOME \ Chemiluminescence Immunoassay (CLIA)	IU/mL	14.2	10%	<5	<100		
(CA242)糖类抗原 242							
Maccura i3000 MAKER \ Chemiluminescence Immunoassay (CLIA)	U/mL	35.4	1.6%	<5	<100		
HYBIOME EVERESYS Series HYBIOME \ Chemiluminescence Immunoassay (CLIA)	U/mL	46.3	10%	<5	<100		
Siemens Immulite1000/2000 Siemens \ Chemiluminescence Immunoassay (CLIA)	U/mL	35.8	10.3%	<5	<100		
Mindray CL-8000i Mindray \ Chemiluminescence Immunoassay (CLIA)	U/mL	246	10%	<5	<100		
(CEA)癌胚抗原							
Abbott Alinity i Series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	60.3	4.7%	10	254		
Beckman UniCel DxI 600/800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	32.8	9.5%	<5	<100		
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	40	9.5%	<5	<100		
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	40.5	9.6%	<5	<100		
Mindray CL-8000i Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	15.2	10%	<5	<100		
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	46.2	11.9%	<5	<100		
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	29.6	9.1%	<5	<100		
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	39.5	11.6%	<5	<100		
HYBIOME EVERESYS Series HYBIOME \ Chemiluminescence Immunoassay (CLIA)	ng/mL	33.6	10%	<5	<100		
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	59.7	12.3%	<5	<100		
(Cyfra 21-1)细胞角蛋白19片段							
Abbott Alinity i Series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	13.9	4%	12	322		
HYBIOME EVERESYS Series HYBIOME \ Chemiluminescence Immunoassay (CLIA)	ng/mL	7.92	10%	<5	<100		
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	15.9	10.7%	<5	<100		
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	11	8.2%	<5	<100		
Maccura i3000 MAKER \ Chemiluminescence Immunoassay (CLIA)	ng/mL	14.9	2.4%	<5	<100		
Mindray CL-8000i Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	25.1	10%	<5	<100		
(Ferritin)铁蛋白							
Mindray CL-8000i Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	337	10%	<5	<100		
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	323	9.9%	<5	<100		
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	329	9.4%	<5	<100		
HYBIOME EVERESYS Series HYBIOME \ Chemiluminescence Immunoassay (CLIA)	ng/mL	408	10%	<5	<100		
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	408	9.6%	<5	<100		
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	342	11.9%	<5	<100		
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	432	11.9%	<5	<100		
Beckman Access Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	289	9.7%	<5	<100		
(F-PSA)游离前列腺特异性抗原							
Abbott Alinity i Series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	8.25	4.6%	12	324		
Mindray CL-8000i Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.14	10%	<5	<100		
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	8.94	11%	<5	<100		

REF	513-2	6x3mL	EXP	2027-11-26	LOT	30F1D64	更新日期:	2024-08
								3 / 4

		批号: 30F1D64			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4.45	9.7%	<5	<100
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	6.45	12.2%	<5	<100
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	4.99	8.6%	<5	<100
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	7.88	12%	<5	<100
HYBIOME EVERESYS Series HYBIOME \ Chemiluminescence Immunoassay (CLIA)	ng/mL	5.57	10%	<5	<100
Beckman Access Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	6.95	9.2%	<5	<100
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	4.9	8.8%	<5	<100
(hCG)人绒毛膜促性腺激素					
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	187	9.6%	<5	<100
Mindray CL-8000i Mindray \ Chemiluminescence Immunoassay (CLIA)	IU/L	110	10%	<5	<100
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/L	126	9.5%	<5	<100
Siemens ADVIA Centaur XP/XPT Dedicated \ Enzyme linked immunosorbent assays (ELISA)	IU/L	122	9%	<5	<100
HYBIOME EVERESYS Series HYBIOME \ Chemiluminescence Immunoassay (CLIA)	IU/L	41.7	10%	<5	<100
(HE4)人附睾蛋白4					
Abbott Alinity i Series Abbott \ Chemiluminescence Immunoassay (CLIA)	pmol/L	253	5.8%	11	276
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	290	9.6%	<5	<100
HYBIOME EVERESYS Series HYBIOME \ Chemiluminescence Immunoassay (CLIA)	pmol/L	144	10%	<5	<100
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	pmol/L	276	9.8%	<5	<100
Mindray CL-8000i Mindray \ Chemiluminescence Immunoassay (CLIA)	pmol/L	41.5	10%	<5	<100
Maccura i3000 MAKER \ Chemiluminescence Immunoassay (CLIA)	pmol/L	257	4.3%	<5	<100
(PSA)前列腺特异性抗原					
Abbott Alinity i Series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	15.2	4.5%	12	314
Mindray CL-8000i Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	10.4	10%	<5	<100
Roche Modular E170 Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	18.9	9.5%	<5	<100
HYBIOME EVERESYS Series HYBIOME \ Chemiluminescence Immunoassay (CLIA)	ng/mL	22.9	10%	<5	<100
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	20.4	9.3%	<5	<100
Siemens ADVIA Centaur XP/XPT Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	18.6	9.7%	<5	<100
Beckman Access Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	19.9	9%	<5	<100
Siemens ADVIA Centaur XP/XPT Siemens \ Chemiluminescence enzyme immunoassay (CLEIA)	ng/mL	18.9	9.5%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	17.9	10.6%	<5	<100
Roche Modular E170 Roche \ Electro-Chemiluminescence Immunoassay (ECLIA)	ng/mL	17.6	10.8%	<5	<100
(SCC)鳞状细胞癌抗原					
Abbott Alinity i Series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	16.8	3.2%	12	298
Abbott ARCHITECT i series Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	16.8	9.5%	<5	<100
HYBIOME EVERESYS Series HYBIOME \ Chemiluminescence Immunoassay (CLIA)	ng/mL	14.7	10%	<5	<100
Mindray CL-8000i Mindray \ Chemiluminescence Immunoassay (CLIA)	ng/mL	10.3	10%	<5	<100
Abbott ARCHITECT i series Abbott \ Chemiluminescence Immunoassay (CLIA)	ng/mL	18.3	10.4%	<5	<100
Maccura i3000 MAKER \ Chemiluminescence Immunoassay (CLIA)	ng/mL	17.1	0.9%	<5	<100
(TPA)组织多肽抗原					
DiaSorin Liaison /XL Diasorin \ Chemiluminescence Immunoassay (CLIA)	U/L	2476	11.8%	<5	<100
DiaSorin Liaison /XL Dedicated \ Chemiluminescence Immunoassay (CLIA)	U/L	7363	9.9%	<5	<100
(TPS)组织多肽特异抗原					
Siemens Immulite1000/2000 Siemens \ Chemiluminescence Immunoassay (CLIA)	U/L	82.8	12%	<5	<100
Siemens Immulite1000/2000 Dedicated \ Chemiluminescence Immunoassay (CLIA)	U/L	605	10.1%	<5	<100

REF	513-2	6x3mL	EXP	2027-11-26	LOT	30F1D64	更新日期:	2024-08
								4 / 4

		批号: 30F1D64			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(β-hCG)人绒毛膜促性腺激素β亚基					
Beckman Access Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	0.31	10%	<5	<100
HYBIOME EVERESYS Series HYBIOME \ Chemiluminescence Immunoassay (CLIA)	IU/mL	0.046	10%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。