

<b>REF</b>	521-1	6x3mL	<b>EXP</b>	2024-10-17	<b>LOT</b>	31B2AEE	更新日期:	2024-08
								1 / 2

		批号: 31B2AEE					
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数			
<b>(Alb)白蛋白</b>							
Roche cobas 4000/6000/8000	g/L	20.3	12%	<5	<100		
Roche \ Bromocresol green							
<b>(ASO)抗链球菌溶血素 O 抗体</b>							
Siemens BN ProSpec	IU/mL	133	3.1%	16	344		
Dedicated \ Immunonephelometry (INA)							
Beckman Immage 800	IU/mL	82.5	12%	<5	<100		
Beckman Coulter \ Immunonephelometry (INA)							
<b>(AT-III)抗凝血酶III</b>							
Stago STA series	%	7	12%	<5	<100		
Dedicated \ Nephelometry							
<b>(C3)补体 C3</b>							
Beckman Immage 800	mg/dL	73.6	8.1%	<5	<100		
Beckman Coulter \ Immunonephelometry (INA)							
Siemens BN II	mg/dL	78.8	4.9%	16	346		
Dedicated \ Immunonephelometry (INA)							
<b>(C4)补体 C4</b>							
Beckman Immage 800	mg/dL	18.5	9.3%	<5	<100		
Beckman Coulter \ Immunonephelometry (INA)							
Siemens BN II	mg/dL	18.7	4.6%	13	273		
Dedicated \ Immunonephelometry (INA)							
<b>(Cer)铜蓝蛋白</b>							
Siemens BN ProSpec	g/L	0.165	2.2%	6	130		
Dedicated \ Immunonephelometry (INA)							
Siemens BN II	g/L	0.169	4%	10	209		
Dedicated \ Immunonephelometry (INA)							
Beckman Immage 800	g/L	0.132	12%	<5	<100		
Beckman Coulter \ Immunonephelometry (INA)							
<b>(CRP)C反应蛋白</b>							
Beckman AU series	mg/dL	1.5	2.2%	7	138		
DongOu \ Immunoturbidimetry (ITA)							
Beckman Immage 800	mg/dL	1.36	12%	<5	<100		
Beckman Coulter \ Immunonephelometry (INA)							
<b>(Ferritin)铁蛋白</b>							
Roche cobas 4000/6000/8000	ng/mL	66.4	12%	<5	<100		
Roche \ Immunoturbidimetry (ITA)							
<b>(IgA)免疫球蛋白 IgA</b>							
Beckman Immage 800	mg/dL	121	8.3%	<5	<100		
Beckman Coulter \ Immunonephelometry (INA)							
Siemens BN II	mg/dL	136	5.9%	16	343		
Dedicated \ Immunonephelometry (INA)							
<b>(IgE)免疫球蛋白 IgE</b>							
Beckman Immage 800	IU/mL	28.7	12.4%	<5	<100		
Beckman Coulter \ Immunonephelometry (INA)							
Siemens BN II	IU/mL	31.3	4.1%	10	212		
Dedicated \ Immunonephelometry (INA)							
<b>(IgG)免疫球蛋白 IgG</b>							
Beckman Immage 800	mg/dL	683	9%	<5	<100		
Beckman Coulter \ Immunonephelometry (INA)							
Siemens BN II	mg/dL	751	4.3%	16	344		
Dedicated \ Immunonephelometry (INA)							
<b>(IgM)免疫球蛋白 IgM</b>							
Beckman Immage 800	mg/dL	83.3	9%	<5	<100		
Beckman Coulter \ Immunonephelometry (INA)							
Siemens BN II	mg/dL	87.6	4.4%	16	342		
Dedicated \ Immunonephelometry (INA)							
<b>(Kappa Light Chain)κ轻链</b>							
Beckman Immage 800	g/L	5.85	12%	<5	<100		
Beckman Coulter \ Immunonephelometry (INA)							
<b>(Lambda Light Chain)λ 轻链</b>							
Beckman Immage 800	g/L	2.97	12%	<5	<100		
Beckman Coulter \ Immunonephelometry (INA)							
<b>(PAB)前白蛋白</b>							
Beckman AU series	g/L	0.165	2.5%	27	564		
DongOu \ Immunoturbidimetry (ITA)							
Roche cobas 4000/6000/8000	g/L	0.161	12%	<5	<100		
DiaSys \ Immunonephelometry (INA)							
<b>(RBP)视黄醇结合蛋白</b>							
Beckman AU series	mg/dL	2.78	3%	27	561		
DongOu \ Immunoturbidimetry (ITA)							
Roche cobas 4000/6000/8000	mg/dL	2.76	12%	<5	<100		
WanTai \ Immunonephelometry (INA)							
<b>(RF)类风湿因子</b>							
Siemens BN ProSpec	IU/mL	38.6	5%	16	346		
Dedicated \ Immunonephelometry (INA)							
Beckman Immage 800	IU/mL	36.1	12%	<5	<100		
Beckman Coulter \ Immunonephelometry (INA)							

<b>REF</b>	521-1	6x3mL	<b>EXP</b>	2024-10-17	<b>LOT</b>	31B2AEE	更新日期:	2024-08
								2 / 2

批号: 31B2AEE					
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
<b>(TRF)转铁蛋白</b>					
Roche cobas 4000/6000/8000 Roche \ Immunoturbidimetry (ITA)	mg/dL	127	12%	<5	<100
<b>(β2M)β-2-微球蛋白</b>					
Beckman AU series LEADMAN \ Immunoturbidimetry (ITA)	ng/mL	2625	2.8%	27	561
Roche cobas 4000/6000/8000 JS (Shanghai) \ Immunonephelometry (INA)	ng/mL	2410	12%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。