

REF	521-2	6x3mL	EXP	2023-11-17	LOT	31B519B	更新日期:	2024-08
								1 / 3

		批号: 31B519B			
项目\仪器\试剂方法	单位	均值\CV		Lab\点数	
(Alb)白蛋白					
Roche Hitachi/Hitachi Systems Roche \ Bromocresol purple	g/L	35.6	10.4%	<5	<100
Beckman AU series Beckman Coulter \ Bromocresol green	g/L	40.3	0.9%	<5	<100
(ASO)抗链球菌溶血素 O 抗体					
Mindray BS series Mindray \ Immunonephelometry (INA)	IU/mL	260	3.2%	8	<100
Beckman Immage 800 Beckman Coulter \ Immunoturbidimetry (ITA)	IU/mL	179	7.3%	8	204
TOSHIBA TBA MedicalSystem \ Immunonephelometry (INA)	IU/mL	285	4%	<5	<100
Beckman Immage 800 Beckman Coulter \ Immunonephelometry (INA)	IU/mL	158	17%	20	473
Roche Hitachi/Hitachi Systems Wako \ Immunoturbidimetry (ITA)	IU/mL	166	10.2%	<5	<100
Beckman Immage 800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	132	9.1%	<5	<100
Beckman AU series Beckman Coulter \ Immunoturbidimetry (ITA)	IU/mL	163	10.4%	<5	<100
(C3)补体 C3					
Beckman Immage 800 Beckman Coulter \ Immunoturbidimetry (ITA)	mg/dL	135	3.8%	8	204
TOSHIBA TBA MedicalSystem \ Immunoturbidimetry (ITA)	mg/dL	150	5.4%	<5	<100
Mindray BS series Mindray \ Immunoturbidimetry (ITA)	mg/dL	167	4.1%	10	103
Beckman Immage 800 Beckman Coulter \ Immunonephelometry (INA)	mg/dL	137	1.1%	<5	<100
Beckman AU series Beckman Coulter \ Immunoturbidimetry (ITA)	mg/dL	132	9.8%	<5	<100
Beckman Immage 800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	mg/dL	132	9.1%	<5	<100
(C4)补体 C4					
Beckman Immage 800 Beckman Coulter \ Immunoturbidimetry (ITA)	mg/dL	35.2	5%	8	204
TOSHIBA TBA MedicalSystem \ Immunoturbidimetry (ITA)	mg/dL	35.5	5.9%	<5	<100
Mindray BS series Mindray \ Immunoturbidimetry (ITA)	mg/dL	38.9	3.3%	10	101
Beckman Immage 800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	mg/dL	32.6	10.4%	<5	<100
Beckman AU series Beckman Coulter \ Immunonephelometry (INA)	mg/dL	35	8.6%	<5	<100
Beckman Immage 800 Beckman Coulter \ Immunonephelometry (INA)	mg/dL	36.2	1.1%	<5	<100
(Cer)铜蓝蛋白					
Siemens BN II Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	g/L	0.326	9.8%	<5	<100
(CRP)C反应蛋白					
Roche Hitachi/Hitachi Systems chongqiong zybio \ Immunoturbidimetry (ITA)	mg/dL	2.76	7.2%	<5	<100
Beckman AU series Beckman Coulter \ Immunoturbidimetry (ITA)	mg/dL	2.53	12.6%	<5	<100
Beckman AU series KHB \ Immunoturbidimetry (ITA)	mg/dL	2.21	9%	<5	<100
Beckman Immage 800 Beckman Coulter \ Immunonephelometry (INA)	mg/dL	2.35	2.6%	<5	<100
Beckman Immage 800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	mg/dL	2.12	10.4%	<5	<100
(Hap (HPT))触珠蛋白 (结合珠蛋白)					
Siemens BN II Dedicated \ Chemiluminescence enzyme immunoassay (CLEIA)	mg/L	1423	9.6%	<5	<100
(IgA)免疫球蛋白 IgA					
Mindray BS series Mindray \ Immunoturbidimetry (ITA)	mg/dL	296	3.5%	9	<100
Beckman Immage 800 Beckman Coulter \ Immunoturbidimetry (ITA)	mg/dL	224	5.1%	8	205
TOSHIBA TBA MedicalSystem \ Immunoturbidimetry (ITA)	mg/dL	225	3.1%	<5	<100
Roche Hitachi/Hitachi Systems Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	mg/dL	215	10.7%	<5	<100
Beckman Immage 800 Beckman Coulter \ Immunonephelometry (INA)	mg/dL	225	2%	<5	<100
Beckman AU series Beckman Coulter \ Immunoturbidimetry (ITA)	mg/dL	219	10.5%	<5	<100
Beckman Immage 800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	mg/dL	221	10.4%	<5	<100
(IgE)免疫球蛋白 IgE					
Beckman Immage 800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	39.5	9.6%	<5	<100

REF	521-2	6x3mL	EXP	2023-11-17	LOT	31B519B	更新日期:	2024-08
								2 / 3

		批号: 31B519B			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
Beckman AU series Beckman Coulter \ Immunoturbidimetry (ITA)	IU/mL	41.3	11.9%	<5	<100
(IgG)免疫球蛋白 IgG					
Beckman Immage 800 Beckman Coulter \ Immunoturbidimetry (ITA)	mg/dL	1281	3.8%	8	205
Beckman Immage 800 Beckman Coulter \ Immunonephelometry (INA)	mg/dL	1314	4.1%	20	482
TOSHIBA TBA MedicalSystem \ Immunoturbidimetry (ITA)	mg/dL	1484	2.2%	<5	<100
Mindray BS series Mindray \ Immunoturbidimetry (ITA)	mg/dL	1745	22.4%	10	<100
Beckman Immage 800 Beckman Coulter \ Chemiluminescence Immunoassay (CLIA)	mg/dL	1245	9.9%	<5	<100
Roche Hitachi/Hitachi Systems Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	mg/dL	1233	10%	<5	<100
Beckman AU series Beckman Coulter \ Immunoturbidimetry (ITA)	mg/dL	1250	10.4%	<5	<100
(IgM)免疫球蛋白 IgM					
Mindray BS series Mindray \ Immunoturbidimetry (ITA)	mg/dL	155	3.8%	9	103
Beckman Immage 800 Beckman Coulter \ Immunoturbidimetry (ITA)	mg/dL	171	4.1%	8	206
Beckman Immage 800 Beckman Coulter \ Immunonephelometry (INA)	mg/dL	12.8	312.9%	20	485
Beckman AU series Beckman Coulter \ Immunoturbidimetry (ITA)	mg/dL	133	16.5%	<5	<100
Roche Hitachi/Hitachi Systems Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	mg/dL	136	11.8%	<5	<100
Beckman Immage 800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	mg/dL	121	13.2%	<5	<100
(Kappa Light Chain)κ轻链					
Siemens BN II Dedicated \ Chemiluminescence Immunoassay (CLIA)	g/L	9.89	9.9%	<5	<100
Beckman Immage 800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	g/L	10.3	10.7%	<5	<100
(Lambda Light Chain)λ 轻链					
Siemens BN II Dedicated \ Chemiluminescence Immunoassay (CLIA)	g/L	4.61	10.4%	<5	<100
Beckman Immage 800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	g/L	5.52	9.2%	<5	<100
(PAB)前白蛋白					
Beckman AU series KINGSBIO \ Immunoturbidimetry (ITA)	g/L	0.318	2.2%	19	521
Beckman Immage 800 Beckman Coulter \ Immunonephelometry (INA)	g/L	0.326	1.5%	<5	<100
Beckman AU series Beckman Coulter \ Immunoturbidimetry (ITA)	g/L	0.287	9.8%	<5	<100
Beckman Immage 800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	g/L	0.362	9.4%	<5	<100
(RBP)视黄醇结合蛋白					
Beckman AU series chongqiong zybio \ Immunoturbidimetry (ITA)	mg/dL	4.66	2.3%	5	138
Beckman Immage 800 Beckman Coulter \ Immunonephelometry (INA)	mg/dL	6.09	2.3%	<5	<100
Siemens BN II Dedicated \ Chemiluminescence Immunoassay (CLIA)	mg/dL	4.46	10.3%	<5	<100
(RF)类风湿因子					
Beckman Immage 800 Beckman Coulter \ Immunonephelometry (INA)	IU/mL	59.4	9.8%	20	478
Mindray BS series Mindray \ Immunoturbidimetry (ITA)	IU/mL	64.5	4.1%	10	<100
Beckman Immage 800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	IU/mL	60.2	9.6%	<5	<100
Beckman AU series Beckman Coulter \ Immunoturbidimetry (ITA)	IU/mL	64.8	9.7%	<5	<100
Roche Hitachi/Hitachi Systems Wako \ Immunoturbidimetry (ITA)	IU/mL	59.6	10.2%	<5	<100
TOSHIBA TBA MedicalSystem \ Immunoturbidimetry (ITA)	IU/mL	88.3	0%	<5	<100
(TRF)转铁蛋白					
Beckman Immage 800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	mg/dL	242	9.5%	<5	<100
Beckman Immage 800 Beckman Coulter \ Immunonephelometry (INA)	mg/dL	228	1.7%	<5	<100
(α-1-AG)α-1-酸性糖蛋白					
Beckman AU series Beckman Coulter \ Immunonephelometry (INA)	g/L	0.884	9.7%	<5	<100
(α-1-AT)α-1-抗胰蛋白酶					
Beckman AU series Beckman Coulter \ Immunoturbidimetry (ITA)	mg/dL	127	11%	<5	<100

REF	521-2	6x3mL	EXP	2023-11-17	LOT	31B519B	更新日期:	2024-08
								3 / 3

		批号: 31B519B			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
(β2M)β-2-微球蛋白					
Mindray BS series Mindray \ OTHER	ng/mL	3989	4%	7	<100
TOSHIBA TBA MedicalSystem \ OTHER	ng/mL	3725	5%	<5	<100
Beckman AU series chongqiong zybio \ Immunoturbidimetry (ITA)	ng/mL	3548	2.8%	19	548
Beckman Immage 800 Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	4400	13.6%	<5	<100
Siemens BN II Dedicated \ Chemiluminescence Immunoassay (CLIA)	ng/mL	3200	12.5%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。