

<b>REF</b>	521-2	6x3mL	<b>EXP</b>	2026-11-06	<b>LOT</b>	31B52C7	更新日期:	2024-08
								1 / 2

		批号: 31B52C7			
项目\仪器\试剂方法	单位	均值 \ CV		Lab \ 点数	
<b>(Alb)白蛋白</b>					
Olympus AU400					
Olympus \ Bromocresol green	g/L	0.039	9.8%	<5	<100
<b>(ASO)抗链球菌溶血素 O 抗体</b>					
Beckman AU series					
BSBE \ OTHER	IU/mL	204	2.5%	5	<100
Abbott ARCHITECT c series					
BSBE \ OTHER	IU/mL	222	0.8%	<5	<100
Siemens ADVIA 1800/2400					
DiaSys \ Immunoturbidimetry (ITA)	IU/mL	246	3.3%	<5	134
Olympus AU400					
Olympus \ Immunonephelometry (INA)	IU/mL	273	9.7%	<5	<100
<b>(AT-III)抗凝血酶III</b>					
Beckman Immage 800					
Beckman Coulter \ Nephelometry	g/L	0.243	9.7%	<5	<100
<b>(C3)补体 C3</b>					
Siemens ADVIA 1800/2400					
DiaSys \ Immunoturbidimetry (ITA)	mg/dL	135	3.2%	6	268
Olympus AU400					
Olympus \ Immunonephelometry (INA)	mg/dL	138	9.7%	<5	<100
<b>(C4)补体 C4</b>					
Siemens ADVIA 1800/2400					
DiaSys \ Immunoturbidimetry (ITA)	mg/dL	32.7	3.6%	6	271
Olympus AU400					
Olympus \ Immunonephelometry (INA)	mg/dL	40.5	9.7%	<5	<100
<b>(Cer)铜蓝蛋白</b>					
Olympus AU400					
Olympus \ Immunonephelometry (INA)	g/L	0.497	9.7%	<5	<100
<b>(CRP)C反应蛋白</b>					
Olympus AU400					
Olympus \ Immunonephelometry (INA)	mg/dL	3.07	9.4%	<5	<100
<b>(Ferritin)铁蛋白</b>					
Olympus AU400					
Olympus \ Immunoturbidimetry (ITA)	ng/mL	139	9.7%	<5	<100
Siemens BN II					
Siemens \ Immunoturbidimetry (ITA)	ng/mL	119	9.7%	<5	<100
<b>(Hap (HPT) )触珠蛋白 (结合珠蛋白)</b>					
Abbott ARCHITECT c series					
chongqiong zybjo \ OTHER	mg/L	1343	3.8%	6	<100
Beckman AU series					
chongqiong zybjo \ OTHER	mg/L	1418	4.4%	5	<100
Abbott ARCHITECT c series					
chongqiong zybjo \ Immunonephelometry (INA)	mg/L	1350	10%	<5	<100
Beckman AU series					
chongqiong zybjo \ Immunonephelometry (INA)	mg/L	1300	10%	<5	<100
Olympus AU400					
Olympus \ Immunonephelometry (INA)	mg/L	1390	9.7%	<5	<100
<b>(IgA )免疫球蛋白 IgA</b>					
Siemens ADVIA 1800/2400					
DiaSys \ Immunoturbidimetry (ITA)	mg/dL	273	5.1%	6	269
Olympus AU400					
Olympus \ Immunonephelometry (INA)	mg/dL	279	9.7%	<5	<100
<b>(IgE)免疫球蛋白 IgE</b>					
Beckman Immage 800					
Beckman Coulter \ Immunonephelometry (INA)	IU/mL	48.5	9.7%	<5	<100
<b>(IgG)免疫球蛋白 IgG</b>					
Siemens ADVIA 1800/2400					
DiaSys \ Immunoturbidimetry (ITA)	mg/dL	1437	4.1%	6	272
Olympus AU400					
Olympus \ Immunonephelometry (INA)	mg/dL	1421	9.7%	<5	<100
<b>(IgM)免疫球蛋白 IgM</b>					
Siemens ADVIA 1800/2400					
DiaSys \ Immunoturbidimetry (ITA)	mg/dL	170	4.7%	6	268
Olympus AU400					
Olympus \ Immunonephelometry (INA)	mg/dL	170	9.7%	<5	<100
<b>(Kappa Light Chain)κ轻链</b>					
Beckman Immage 800					
Beckman Coulter \ Immunonephelometry (INA)	g/L	11	9.7%	<5	<100
<b>(Lambda Light Chain)λ 轻链</b>					
Beckman Immage 800					
Beckman Coulter \ Immunonephelometry (INA)	g/L	5.85	9.7%	<5	<100
<b>(PAB)前白蛋白</b>					
Olympus AU400					
Olympus \ Immunonephelometry (INA)	g/L	0.324	9.7%	<5	<100
<b>(RBP)视黄醇结合蛋白</b>					
Siemens ADVIA 1800/2400					
BIOSINO \ Immunoturbidimetry (ITA)	mg/dL	5.56	4.1%	6	273

<b>REF</b>	521-2	6x3mL	<b>EXP</b>	2026-11-06	<b>LOT</b>	31B52C7	更新日期:	2024-08
								2 / 2

		批号: 31B52C7			
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数	
Siemens BN II Siemens \ Immunonephelometry (INA)	mg/dL	4.87	9.7%	<5	<100
<b>(RF)类风湿因子</b>					
Siemens ADVIA 1800/2400 DiaSys \ Immunoturbidimetry (ITA)	IU/mL	61.5	5.2%	6	271
Abbott ARCHITECT c series BSBE \ OTHER	IU/mL	72.5	0.9%	<5	<100
Beckman Immage 800 Beckman Coulter \ Immunonephelometry (INA)	IU/mL	69.1	9.7%	<5	<100
<b>(TRF)转铁蛋白</b>					
Siemens ADVIA 1800/2400 DiaSys \ Immunoturbidimetry (ITA)	mg/dL	231	1.5%	<5	<100
Olympus AU400 Olympus \ Immunoturbidimetry (ITA)	mg/dL	242	9.7%	<5	<100
<b>(α-1-AG)α-1-酸性糖蛋白</b>					
Olympus AU400 Olympus \ Immunoturbidimetry (ITA)	g/L	0.795	9.7%	<5	<100
<b>(α-1-AT)α-1-抗胰蛋白酶</b>					
Olympus AU400 Olympus \ Immunoturbidimetry (ITA)	mg/dL	153	9.7%	<5	<100
<b>(α2M)α-2-巨球蛋白</b>					
Beckman Immage 800 Beckman Coulter \ Immunonephelometry (INA)	mg/L	2250	9.7%	<5	<100
<b>(β2M)β-2-微球蛋白</b>					
Abbott ARCHITECT c series BSBE \ OTHER	ng/mL	3721	4.1%	6	<100
Beckman AU series BSBE \ OTHER	ng/mL	3422	5.9%	<5	<100
Olympus AU400 Olympus \ Immunonephelometry (INA)	ng/mL	4400	9.1%	<5	<100

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。