

REF	521-3	6x3mL	EXP	2025-12-22	LOT	31B79CD	更新日期:	2024-08
								1 / 2

		批号: 31B79CD					
项目\仪器\试剂\方法	单位	均值\CV		Lab\点数			
(Alb)白蛋白							
Olympus AU400							
Olympus \ Bromocresol green	g/L	58	10%	<5	<100		
(ASO)抗链球菌溶血素 O 抗体							
Abbott ARCHITECT c series							
BSBE \ OTHER	IU/mL	324	1.3%	7	192		
Siemens BN ProSpec							
Dedicated \ Immunonephelometry (INA)	IU/mL	348	2.9%	<5	<100		
Siemens BN II							
Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	343	2.9%	<5	<100		
Olympus AU400							
Olympus \ Immunonephelometry (INA)	IU/mL	344	10%	<5	<100		
(AT-III)抗凝血酶III							
Beckman Immage 800							
Beckman Coulter \ Nephelometry	g/L	0.372	10%	<5	<100		
(C3)补体 C3							
Siemens BN II							
Dedicated \ Immunonephelometry (INA)	mg/dL	221	4.5%	<5	<100		
Siemens BN II							
Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	mg/dL	250	4.3%	<5	<100		
Olympus AU400							
Olympus \ Immunonephelometry (INA)	mg/dL	202	10%	<5	<100		
(C4)补体 C4							
Siemens BN II							
Dedicated \ Immunonephelometry (INA)	mg/dL	57.5	6.7%	<5	<100		
Siemens BN II							
Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	mg/dL	64.9	6.6%	<5	<100		
Olympus AU400							
Olympus \ Immunonephelometry (INA)	mg/dL	57.3	10%	<5	<100		
(Cer)铜蓝蛋白							
Siemens BN II							
Dedicated \ Immunonephelometry (INA)	g/L	0.68	3.5%	<5	<100		
Olympus AU400							
Olympus \ Immunonephelometry (INA)	g/L	0.669	10%	<5	<100		
(CRP)C反应蛋白							
Siemens BN II							
Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	mg/dL	6.5	3.7%	<5	<100		
Beckman AU series							
DongOu \ Immunoturbidimetry (ITA)	mg/dL	6.93	1.8%	<5	<100		
Olympus AU400							
Olympus \ Immunonephelometry (INA)	mg/dL	6.99	10%	<5	<100		
(Ferritin)铁蛋白							
Olympus AU400							
Olympus \ Immunoturbidimetry (ITA)	ng/mL	340	10%	<5	<100		
Siemens BN II							
Siemens \ Immunoturbidimetry (ITA)	ng/mL	304	10%	<5	<100		
(Hap (HPT))触珠蛋白 (结合珠蛋白)							
Abbott ARCHITECT c series							
BSBE \ OTHER	mg/L	1946	4.3%	<5	<100		
Abbott ARCHITECT c series							
chongqiong zybio \ OTHER	mg/L	2244	6.8%	11	260		
Olympus AU400							
Olympus \ Enzyme linked immunosorbent assays (ELISA)	mg/L	2020	10%	<5	<100		
(IgA)免疫球蛋白 IgA							
Siemens BN II							
Dedicated \ Immunonephelometry (INA)	mg/dL	470	5.7%	<5	<100		
Siemens BN II							
Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	mg/dL	534	4.6%	<5	<100		
Olympus AU400							
Olympus \ Immunonephelometry (INA)	mg/dL	395	10%	<5	<100		
(IgE)免疫球蛋白 IgE							
Siemens BN II							
Dedicated \ Immunonephelometry (INA)	IU/mL	63.3	3.1%	<5	<100		
Siemens BN II							
Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	63.4	2.8%	<5	<100		
Beckman Immage 800							
Beckman Coulter \ Immunonephelometry (INA)	IU/mL	41.9	10%	<5	<100		
(IgG)免疫球蛋白 IgG							
Siemens BN II							
Dedicated \ Immunonephelometry (INA)	mg/dL	2220	4.5%	<5	<100		
Siemens BN II							
Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	mg/dL	2304	6.4%	<5	<100		
Olympus AU400							
Olympus \ Immunonephelometry (INA)	mg/dL	1997	10%	<5	<100		

REF	521-3	6x3mL	EXP	2025-12-22	LOT	31B79CD	更新日期：	2024-08
								2 / 2

批号： 31B79CD						
项目\仪器\试剂\方法	单位	均值 \ CV		Lab \ 点数		
(IgM)免疫球蛋白 IgM						
Siemens BN II Dedicated \ Immunonephelometry (INA)	mg/dL	280	3.9%	<5	<100	
Siemens BN II Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	mg/dL	300	3.9%	<5	<100	
Olympus AU400 Olympus \ Immunonephelometry (INA)	mg/dL	225	10%	<5	<100	
(Kappa Light Chain)κ轻链						
Beckman Immage 800 Beckman Coulter \ Immunonephelometry (INA)	g/L	16.4	10%	<5	<100	
(Lambda Light Chain)λ 轻链						
Beckman Immage 800 Beckman Coulter \ Immunonephelometry (INA)	g/L	9.03	10%	<5	<100	
(PAB)前白蛋白						
Beckman AU series DongOu \ Immunoturbidimetry (ITA)	g/L	0.545	2.7%	12	241	
Olympus AU400 Olympus \ Immunonephelometry (INA)	g/L	0.539	10%	<5	<100	
(RBP)视黄醇结合蛋白						
Beckman AU series DongOu \ Immunoturbidimetry (ITA)	mg/dL	8.21	2.1%	12	236	
Siemens BN II Siemens \ Immunonephelometry (INA)	mg/dL	9.13	10%	<5	<100	
(RF)类风湿因子						
Siemens BN ProSpec Dedicated \ Immunonephelometry (INA)	IU/mL	85.1	2.1%	<5	<100	
Siemens BN II Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	IU/mL	83.6	5.2%	<5	<100	
Beckman Immage 800 Beckman Coulter \ Immunonephelometry (INA)	IU/mL	108	10%	<5	<100	
(TRF)转铁蛋白						
Siemens BN II Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	mg/dL	414	6.6%	<5	<100	
Abbott ARCHITECT c series BSBE \ OTHER	mg/dL	416	2.2%	<5	<100	
Olympus AU400 Olympus \ Immunonephelometry (INA)	mg/dL	374	10%	<5	<100	
(α-1-AG)α-1-酸性糖蛋白						
Siemens BN II Dedicated \ Electro-Chemiluminescence Immunoassay (ECLIA)	g/L	1.67	10.2%	<5	<100	
Olympus AU400 Olympus \ Enzyme linked immunosorbent assays (ELISA)	g/L	1.24	10%	<5	<100	
(α-1-AT)α-1-抗胰蛋白酶						
Beckman AU series OTHER \ Immunoturbidimetry (ITA)	mg/dL	241	3%	12	242	
Olympus AU400 Olympus \ Immunoturbidimetry (ITA)	mg/dL	223	10%	<5	<100	
(α2M)α-2-巨球蛋白						
Beckman Immage 800 Beckman Coulter \ Immunoturbidimetry (ITA)	mg/L	3460	10%	<5	<100	
(β2M)β-2-微球蛋白						
Abbott ARCHITECT c series BSBE \ OTHER	ng/mL	5448	6.9%	15	406	
Beckman AU series LEADMAN \ Immunoturbidimetry (ITA)	ng/mL	6652	2.4%	12	240	
Olympus AU400 Olympus \ Immunonephelometry (INA)	ng/mL	6420	9.3%	<5	<100	

注：此报告所提供的数据均基于检测相同批号质控品的若干实验室的数据汇总统计而来，仅供学习、参考之用。因所用技术、仪器和试剂的不同，或因制造商检测方法的改变，均可导致实验室实际测得的数据偏离此报告所提供的数据。根据良好实验室规范的要求，实验室须遵循相关技术规范确立自己的均值和可接受范围。